

SVENSK STANDARD

SS-EN ISO 11990-2:2014

Fastställt/Approved: 2014-11-07
Publicerad/Published: 2014-11-10
Utgåva/Edition: 2
Språk/Language: engelska/English
ICS: 11.040.10; 31.260

Lasrar och lasertillbehör – Bestämning av resistensen hos trakealtuber mot påverkan av laserstrålning – Del 2: Trakealtubskuffar (ISO 11990-2:2010)

Lasers and laser-related equipment – Determination of laser resistance of tracheal tubes – Part 2: Tracheal tube cuffs (ISO 11990-2:2010)



Standarder får världen att fungera

SIS (Swedish Standards Institute) är en fristående ideell förening med medlemmar från både privat och offentlig sektor. Vi är en del av det europeiska och globala nätverk som utarbetar internationella standarder. Standarder är dokumenterad kunskap utvecklad av framstående aktörer inom industri, näringsliv och samhälle och befrämjar handel över gränser, bidrar till att processer och produkter blir säkrare samt effektiviserar din verksamhet.

Delta och påverka

Som medlem i SIS har du möjlighet att påverka framtida standarder inom ditt område på nationell, europeisk och global nivå. Du får samtidigt tillgång till tidig information om utvecklingen inom din bransch.

Ta del av det färdiga arbetet

Vi erbjuder våra kunder allt som rör standarder och deras tillämpning. Hos oss kan du köpa alla publikationer du behöver – allt från enskilda standarder, tekniska rapporter och standardpaket till handböcker och onlinetjänster. Genom vår webbtjänst e-nav får du tillgång till ett lättnavigerat bibliotek där alla standarder som är aktuella för ditt företag finns tillgängliga. Standarder och handböcker är källor till kunskap. Vi säljer dem.

Utveckla din kompetens och lyckas bättre i ditt arbete

Hos SIS kan du gå öppna eller företagsinterna utbildningar kring innehåll och tillämpning av standarder. Genom vår närhet till den internationella utvecklingen och ISO får du rätt kunskap i rätt tid, direkt från källan. Med vår kunskap om standarders möjligheter hjälper vi våra kunder att skapa verklig nytta och lönsamhet i sina verksamheter.

Vill du veta mer om SIS eller hur standarder kan effektivisera din verksamhet är du välkommen in på www.sis.se eller ta kontakt med oss på tel 08-555 523 00.



Standards make the world go round

SIS (Swedish Standards Institute) is an independent non-profit organisation with members from both the private and public sectors. We are part of the European and global network that draws up international standards. Standards consist of documented knowledge developed by prominent actors within the industry, business world and society. They promote cross-border trade, they help to make processes and products safer and they streamline your organisation.

Take part and have influence

As a member of SIS you will have the possibility to participate in standardization activities on national, European and global level. The membership in SIS will give you the opportunity to influence future standards and gain access to early stage information about developments within your field.

Get to know the finished work

We offer our customers everything in connection with standards and their application. You can purchase all the publications you need from us - everything from individual standards, technical reports and standard packages through to manuals and online services. Our web service e-nav gives you access to an easy-to-navigate library where all standards that are relevant to your company are available. Standards and manuals are sources of knowledge. We sell them.

Increase understanding and improve perception

With SIS you can undergo either shared or in-house training in the content and application of standards. Thanks to our proximity to international development and ISO you receive the right knowledge at the right time, direct from the source. With our knowledge about the potential of standards, we assist our customers in creating tangible benefit and profitability in their organisations.

If you want to know more about SIS, or how standards can streamline your organisation, please visit www.sis.se or contact us on phone +46 (0)8-555 523 00



Europastandarden EN ISO 11990-2:2014 gäller som svensk standard. Detta dokument innehåller den officiella engelska versionen av EN ISO 11990-2:2014.

Denna standard ersätter SS-EN ISO 11990-2:2010, utgåva 1.

The European Standard EN ISO 11990-2:2014 has the status of a Swedish Standard. This document contains the official version of EN ISO 11990-2:2014.

This standard supersedes the Swedish Standard SS-EN ISO 11990-2:2010, edition 1.

**Förhållandet till övriga delar under samma huvudtitel - Utdrag ur Förord i ISO 11990-2:2011/
Relations to other parts under the same general title - Extract from the Foreword of ISO 11990-2:2011**

ISO 11990 consists of the following parts, under the general title *Lasers and laser-related equipment — Determination of laser resistance of tracheal tubes*:

- Part 1: *Tracheal tube shaft*
- Part 2: *Tracheal tube cuffs*

© Copyright/Upphovsrätten till denna produkt tillhör SIS, Swedish Standards Institute, Stockholm, Sverige. Användningen av denna produkt regleras av slutanvändarlicensen som återfinns i denna produkt, se standardens sista sidor.

© Copyright SIS, Swedish Standards Institute, Stockholm, Sweden. All rights reserved. The use of this product is governed by the end-user licence for this product. You will find the licence in the end of this document.

Upplysningar om sakinnehållet i standarden lämnas av SIS, Swedish Standards Institute, telefon 08-555 520 00. Standarder kan beställas hos SIS Förlag AB som även lämnar allmänna upplysningar om svensk och utländsk standard.

Information about the content of the standard is available from the Swedish Standards Institute (SIS), telephone +46 8 555 520 00. Standards may be ordered from SIS Förlag AB, who can also provide general information about Swedish and foreign standards.

Har du synpunkter på innehållet i den här standarden, vill du delta i ett kommande revideringsarbete eller vara med och ta fram andra standarder inom området? Gå in på www.sis.se - där hittar du mer information.

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN ISO 11990-2

October 2014

ICS 11.040.10; 31.260

Supersedes EN ISO 11990-2:2010

English Version

**Lasers and laser-related equipment - Determination of laser
resistance of tracheal tubes - Part 2: Tracheal tube cuffs (ISO
11990-2:2010)**

Lasers et équipements associés aux lasers - Détermination
de la résistance au laser des tubes trachéaux - Partie 2:
Ballonnet de tubes trachéaux (ISO 11990-2:2010)

Laser und Laseranlagen - Bestimmung der Laserresistenz
von Trachealtuben - Teil 2: Trachealtubumanschetten (ISO
11990-2:2010)

This European Standard was approved by CEN on 22 October 2014.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents		Page
Foreword		iv
Introduction		v
1 Scope		1
2 Normative references		1
3 Terms and definitions		1
4 Principle		2
5 Significance and use of the test		2
6 Apparatus		3
6.1 Gas supply system		3
6.2 Containment box		4
6.3 Smoke evacuation device		7
6.4 Lasers and delivery systems		7
6.5 Oxygen analyser		7
7 Reagents and materials		8
8 Preparation of test specimens		8
9 Preparation of apparatus		8
10 Test procedure		9
11 Interpretation of results		10
12 Test report		10
Annex ZA (informative) Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EEC (Medical Devices)		11
Bibliography		12

Foreword

The text of ISO 11990-2:2010 has been prepared by Technical Committee ISO/TC 172 "Optics and photonics" of the International Organization for Standardization (ISO) and has been taken over as EN ISO 11990-2:2014 by Technical Committee CEN/TC 123 "Lasers and photonics" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by April 2015, and conflicting national standards shall be withdrawn at the latest by April 2015.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 11990-2:2010.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive.

For relationship with EU Directive, see informative Annex ZA, which is an integral part of this document.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 11990-2:2010 has been approved by CEN as EN ISO 11990-2:2014 without any modification.

Introduction

A fire in the airway is always a serious matter. In addition to local damage in the larynx, injury can occur to the lower airway and the parenchymal tissue in the lung. The products of combustion may be blown into the lungs.

Procedures performed in the airway, where a tracheal tube and a laser are used, bring together an oxygen-enriched atmosphere, a fuel and high power, the three ingredients necessary to create a fire. The likelihood that a laser beam will contact the tracheal tube during airway procedures is high. This led to the development of a test method, described in ISO 11990-1, to assist the clinician in determining which tracheal tube shaft was the most laser-resistant under a defined set of conditions.

Unfortunately, fires with tracheal tubes, whose shafts were laser-resistant according to ISO 11990-1 have continued to occur. Investigations have shown that the cuff, and not the shaft, of the tracheal tube is the area of lowest laser resistance and most likely to be contacted by the laser beam, even when used according to the manufacturer's instructions. Clinical experience has shown that not only perforation of the part of the shaft below the cuff has happened, but also ignition of the outer surface of the cuff. This could then ignite other parts of the tracheal tube, such as the tip, which is normally unprotected.

Lasers and laser-related equipment — Determination of laser resistance of tracheal tubes —

Part 2: Tracheal tube cuffs

1 Scope

This part of ISO 11990 specifies a method of testing the continuous wave (cw) resistance of the cuff regions of tracheal tubes designed to resist ignition by a laser. Other components of the system, such as the inflation system and shaft (as defined in ISO 11990-1), are outside the scope of this part of ISO 11990.

NOTE 1 The method for testing the laser resistance of the tracheal tube shaft is in the scope of ISO 11990-1.

The specified test method can be used to measure and describe the properties of materials, products or assemblies in response to heat and flame under controlled laboratory conditions. It does not describe or appraise the fire hazard or fire risk of materials, products or assemblies under actual clinical use conditions. However, the results of this test method may be used as an element of a fire risk assessment which takes into account all of the factors that are pertinent to an assessment of the hazard of a particular end use.

NOTE 2 Caution should be observed in interpreting these results, since the direct applicability of the results of this test method to the clinical situation has not been fully established.

NOTE 3 This test method might involve hazardous materials, operations and equipment. This part of ISO 11990 provides advice on minimizing some of the risks associated with its use but does not purport to address all such risks. It is the responsibility of the user of this test method to establish appropriate safety and health practices and determine the applicability of regulatory limitations prior to use.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 11146-1, *Lasers and laser-related equipment — Test methods for laser beam widths, divergence angles and beam propagation ratios — Part 1: Stigmatic and simple astigmatic beams*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1

beam cross-sectional area

A_{95}

smallest area containing 95 % of the total beam power

3.2
beam diameter
 d_{95}
diameter of an aperture in a plane perpendicular to the beam axis which contains 95 % of the total beam power

NOTE Adapted from ISO 11145:2006.

3.3
combustion
any continuing burning process that occurs in or on the test specimen caused by a chemical process of oxidation with the liberation of heat

EXAMPLE Flame, smouldering, rapid evolution of smoke.

3.4
cuff
inflatable balloon permanently attached around the tracheal tube near the patient end to provide an effective seal between the tube and the trachea

[ISO 5361:1999, definition 3.3]

3.5
damage
any change, other than combustion, which may affect the safety of the patient or efficacy of the tracheal tube due to increasing the risk of ignition

EXAMPLE Local heating, melting, creation of holes, pyrolysis.

3.6
ignition
creation of combustion induced by the delivery of power

3.7
laser resistance
measure of the ability of a material to withstand laser power without ignition or damage

4 Principle

WARNING — This test method can result in a rocket-like fire involving the tracheal tube. Such a fire can produce intense heat and light and toxic gases.

To simulate worst-case conditions, the cuff of a tracheal tube is exposed to laser power of known characteristics in an environment of 98 % \pm 2 % oxygen.

5 Significance and use of the test

5.1 This part of ISO 11990 describes a uniform and repeatable test method for measuring the laser resistance of the cuff of a tracheal tube. Most of the variables involved in laser ignition of a tracheal tube have been fixed in order to establish a basis for comparison. This test method for measuring can be used to compare tracheal tubes having differing types and designs of laser protection.

5.2 A large number and range of variables are involved in ignition of a tracheal tube cuff. A change in one variable may affect the outcome of the test. Caution should be observed, since the direct applicability of the results of this test method to the clinical situation has not been fully established.

5.3 Since an oxygen-enriched atmosphere is often present in the clinical situation, either intentionally or unintentionally, the test is performed in an environment of $98 \% \pm 2 \%$ oxygen.

5.4 A flow rate of 1 l/min of oxygen in a 6,0 mm inner diameter tube was chosen as the most appropriate conditions for cuff ignition and establishment of a fire based on studies detailed in the work of Sidebotham, Wolf *et al.* [8].

5.5 The preparation of the cuff of the test specimen shall be in accordance with the manufacturer's instructions for use.

5.6 The majority of manufacturers of laser-resistance cuffs recommend using isotonic saline or water to fill the cuff. For preliminary testing of leakage of the cuff, filling with air is recommended by most manufacturers. This can cause an air bubble, which, in a typical position of the patient during surgery, is not on the top of the filled cuff, but at the area where the cuff and shaft meet. The test report shall include whether a bubble occurs and, if so, report if the bubble fills out the space between the cuff and the underlying shaft material, and whether the shaft material in the cuff region is laser-resistant or not.

NOTE 1 This method can be applied to study the effect of changing the test conditions, but this is outside the scope of this part of ISO 11990. For example, variation of the breathing-gas flow rate or different breathing-gas mixtures might affect the laser resistance of the cuff of a tracheal tube.

NOTE 2 Use of beam cross-sectional shape, other than circular, or mode of laser power delivery, other than continuous wave, may affect the cuff ignition characteristics. Also, cuffs of different construction have different laser resistances.

6 Apparatus

6.1 Gas supply system

6.1.1 The gas supply system shall provide oxygen to the tracheal tube at a controllable flow rate. Also, the system shall be capable of rapidly flooding the containment box with nitrogen or other inert gas or stopping oxygen flow, or both, to extinguish any burning material. An oxygen flow meter and controller and a quick-action inert gas valve shall be part of this system (see Figure 1). The nitrogen or inert gas supplied shall be at a higher pressure and allow a flow rate of at least an order of magnitude greater than that of the oxygen supplied to the tracheal tube.

6.1.2 Other arrangements, such as an oxygen flood valve for rapidly purging the containment box or an inert gas flooding system for rapid extinguishment of burning material, may be used as long as the requirements of the test method as defined herein are not affected.