



Handläggande organ

Fastställt

Utgåva

Sida

Hälsa- och sjukvårdsstandardiseringen, HSS

1999-10-29

1

1 (1+13)

© Copyright SIS. Reproduction in any form without permission is prohibited.

In vitro diagnostic medical devices – Culture media for microbiology – Performance criteria for culture media

The European Standard EN 12322:1999 has the status of a Swedish Standard. This document contains the official English version of EN 12322:1999.

Swedish Standards corresponding to documents referred to in this Standard are listed in "Catalogue of Swedish Standards", issued by SIS. The Catalogue lists, with reference number and year of Swedish approval, International and European Standards approved as Swedish Standards as well as other Swedish Standards.

In vitro-diagnostika – Odlingsmedier för mikrobiologi – Kvalitetskriterier för odlingsmedier

Europastandarden EN 12322:1999 gäller som svensk standard. Detta dokument innehåller den officiella engelska versionen av EN 12322:1999.

Motsvarigheten och aktualiteten i svensk standard till de publikationer som omnämns i denna standard framgår av "Katalog över svensk standard", som ges ut av SIS. I katalogen redovisas internationella och europeiska standarder som fastställts som svenska standarder och övriga gällande svenska standarder.

ICS 07.100.10

Standarder kan beställas hos SIS Förlag AB som även lämnar allmänna upplysningar om svensk och utländsk standard. *Postadress:* SIS, Box 6455, 113 82 STOCKHOLM
Telefon: 08 - 610 30 00. *Telefax:* 08 - 30 77 57

Upplysningar om **sakinnehållet** i standarden lämnas av HSS.
Telefon: 08 - 702 49 00. *Telefax:* 08 - 702 49 15
E-post: hss@hss.se
Prisgrupp P

Tryckt i december 1999

EUROPEAN STANDARD
NORME EUROPÉENNE
EUROPÄISCHE NORM

EN 12322

April 1999

ICS 07.100.10

English version

In vitro diagnostic medical devices - Culture media for microbiology - Performance criteria for culture media

Dispositifs médicaux de diagnostic in vitro - Milieux de culture de microbiologie - Critères de performance des milieux de culture

In-vitro-Diagnostika - Kulturmedien für die Mikrobiologie - Leistungskriterien für Kulturmedien

This European Standard was approved by CEN on 16 March 1999.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the Central Secretariat or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the Central Secretariat has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Central Secretariat: rue de Stassart, 36 B-1050 Brussels

Contents

Foreword	3
Introduction	4
1 Scope	5
2 Normative references	5
3 Definitions	5
4 Performance evaluation	6
4.1 General quality criteria.....	6
4.2 Control strains	7
4.3 Microbiological quality criteria.....	7
4.4 Performance evaluation and interpretation of results.....	9
5 Information to be supplied by the manufacturer.....	9
6 Documentation	9
Annex A (informative) Guidance on preservation and maintenance of control strains	10
Annex B (informative) Bibliography	12

Foreword

This European Standard has been prepared by Technical Committee CEN/TC 140 "In vitro diagnostic medical devices", the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by October 1999, and conflicting national standards shall be withdrawn at the latest by October 1999.

This European Standard has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

Annexes A and B are for information only.

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Czech Republic, Denmark, Finland, France, Germany, Greece, Iceland, Ireland, Italy, Luxembourg, Netherlands, Norway, Portugal, Spain, Sweden, Switzerland and the United Kingdom.

Introduction

In the microbiology laboratory many tests and procedures depend on culture media being consistent and providing reproducible results. Several hundreds of formulae of dehydrated culture media are commercially available, and many more, designed for specific growth responses or purposes are described in the literature. In addition, in clinical and industrial laboratories, the main objectives are growth and rapid and sensitive detection of microorganisms. The requirements for media are specific to both the sample and the organism(s) to be detected. Standardized culture media are therefore a prerequisite for any reliable microbiological work (see annex B [1]), and it is of primary importance to define the major objectives and criteria related to how a culture medium is supposed to perform. Determining the performance characteristics of culture media is necessary for commercially prepared culture media. For in-house prepared culture media the internal quality control is carried out by the user.

Performance criteria for culture media are necessary for obtaining comparable products of the same medium type, regardless of source. In addition these criteria may be used by all microbiology laboratories for the evaluation of nutritive and/or selective properties of (new) culture media.

Uniform performance criteria should therefore result in the manufacture of "standardized" products, and limit the amount of testing of commercial culture media in microbiology laboratories.

1 Scope

This European Standard specifies requirements for the performance of culture media. It is concerned with the traceability, comparability, reproducibility and suitability of culture media used in microbiological laboratories. These characteristics are achieved by applying the quality criteria outlined in this standard.

This European Standard is applicable to :

- a) commercial organizations distributing media to microbiology laboratories in ready-to-use form, as dehydrated media or as semi-finished media (see 2.5 in EN 1659 : 1996) ;
- b) non commercial organizations that distribute media to satellite locations ;
- c) laboratories that prepare culture media for their own use.

Cell culture media are not covered by this standard.

2 Normative references

This European Standard incorporates by dated or undated reference, provisions from other publications. These normative references are cited at the appropriate places in the text and the publications are listed hereafter. For dated references, subsequent amendments to or revisions of any of these publications apply to this European Standard only when incorporated in it by amendment or revision. For undated references the latest edition of the publication referred to applies.

EN 1659 : 1996 In vitro diagnostic systems – Culture media for microbiology – Terms and definitions

EN ISO 8402 : 1995 Quality management and quality assurance – Vocabulary (ISO 8402:1994)

3 Definitions

For the purposes of this standard, the definitions given in EN 1659:1996 and in EN ISO 8402:1995 apply, together with the following :

3.1

batch of culture media; lot of culture media

Fully traceable unit referring to a defined amount of bulk, semi-finished product or end product, which is consistent in type and quality and which has passed the requirements of production (in-process control) and quality assurance testing, and which has been produced within one defined production period having been assigned the same lot number.

3.2

control strain

Microorganism used for microbial performance evaluation of culture media.

3.3 reference strain

Microorganism defined to at least the genus and species level, catalogued and described according to its characteristics.

3.4 reference stock

Lot of containers obtained in the laboratory by a single propagation from a reference strain or multiple containers from the same lot of a reference strain from a supplier.

3.5 stock culture

Subculture(s) of a reference stock.

3.6 working culture

Subculture of a stock culture.

4 Performance evaluation

4.1 General quality criteria

NOTE : This standard is not intended to give a detailed description of quality management practices for manufacturing of culture media, which are described elsewhere (see annex B, [1] to [7], [11] and [13]).

It should be noted that, according to some quality systems in use by manufacturers (e.g. according to EN ISO 9001 and EN 46001 or EN ISO 9002 and EN 46002), besides microbiological performance evaluation of culture media, performance testing shall also include testing of physical and chemical properties (see annex B, [13]), such as :

- a) quantity filled ;
- b) layer thickness and filling format ;
- c) colour ;
- d) clarity and/or optical artifacts ;
- e) homogeneity ;
- f) gel stability and consistency ;
- g) moisture content or homogeneity of dehydrated media ;
- h) appearance of specific media (e.g., blood agar) ;
- i) pH value.