

# SVENSK STANDARD

## SS-EN 14133:2009

Fastställt/Approved: 2009-06-05  
Publicerad/Published: 2009-08-18  
Utgåva/Edition: 2  
Språk/Language: engelska/English  
ICS: 67.160.10

---

### **Foodstuffs – Determination of ochratoxin A in wine and beer – HPLC method with immunoaffinity column clean-up**



SWEDISH  
STANDARDS  
INSTITUTE

# Hitta rätt produkt och ett leveranssätt som passar dig

## Standarder

Genom att följa gällande standard både effektiviserar och säkrar du ditt arbete. Många standarder ingår dessutom ofta i paket.

## Tjänster

Abonnemang är tjänsten där vi uppdaterar dig med aktuella standarder när förändringar sker på dem du valt att abonnera på.

På så sätt är du säker på att du alltid arbetar efter rätt utgåva.

e-nav är vår online-tjänst som ger dig och dina kollegor tillgång till standarder ni valt att abonnera på dygnet runt. Med e-nav kan samma standard användas av flera personer samtidigt.

## Leveranssätt

Du väljer hur du vill ha dina standarder levererade. Vi kan erbjuda dig dem på papper och som pdf.

## Andra produkter

Vi har böcker som underlättar arbetet att följa en standard. Med våra böcker får du ökad förståelse för hur standarder ska följas och vilka fördelar den ger dig i ditt arbete. Vi tar fram många egna publikationer och fungerar även som återförsäljare. Det gör att du hos oss kan hitta över 500 unika titlar. Vi har även tekniska rapporter, specifikationer och "workshop agreement".

Matriser är en översikt på standarder och handböcker som bör läsas tillsammans. De finns på sis.se och ger dig en bra bild över hur olika produkter hör ihop.

## Standardiseringsprojekt

Du kan påverka innehållet i framtida standarder genom att delta i någon av SIS ca 400 Tekniska Kommittéer.

# Find the right product and the type of delivery that suits you

## Standards

By complying with current standards, you can make your work more efficient and ensure reliability. Also, several of the standards are often supplied in packages.

## Services

Subscription is the service that keeps you up to date with current standards when changes occur in the ones you have chosen to subscribe to. This ensures that you are always working with the right edition.

e-nav is our online service that gives you and your colleagues access to the standards you subscribe to 24 hours a day. With e-nav, the same standards can be used by several people at once.

## Type of delivery

You choose how you want your standards delivered. We can supply them both on paper and as PDF files.

## Other products

We have books that facilitate standards compliance. They make it easier to understand how compliance works and how this benefits you in your operation. We produce many publications of our own, and also act as retailers. This means that we have more than 500 unique titles for you to choose from. We also have technical reports, specifications and workshop agreements.

Matrices, listed at sis.se, provide an overview of which publications belong together.

## Standardisation project

You can influence the content of future standards by taking part in one or other of SIS's 400 or so Technical Committees.

Europastandarden EN 14133:2009 gäller som svensk standard. Detta dokument innehåller den officiella engelska versionen av EN 14133:2009.

Denna standard ersätter SS-EN 14133:2004, utgåva 1; SS-EN 14133:2004/AC:2007, utgåva 1.

The European Standard EN 14133:2009 has the status of a Swedish Standard. This document contains the official English version of EN 14133:2009.

This standard supersedes the Swedish Standard SS-EN 14133:2004, edition 1;  
SS-EN 14133:2004/AC:2007, edition 1;

© Copyright/Upphovsrätten till denna produkt tillhör SIS, Swedish Standards Institute, Stockholm, Sverige. Användningen av denna produkt regleras av slutanvändarlicensen som återfinns i denna produkt, se standardens sista sidor.

© Copyright SIS, Swedish Standards Institute, Stockholm, Sweden. All rights reserved. The use of this product is governed by the end-user licence for this product. You will find the licence in the end of this document.

Upplysningar om sakinnehållet i standarden lämnas av SIS, Swedish Standards Institute, telefon 08-555 520 00.

Standarder kan beställas hos SIS Förlag AB som även lämnar allmänna upplysningar om svensk och utländsk standard.

Information about the content of the standard is available from the Swedish Standards Institute (SIS), tel +46 8 555 520 00.

Standards may be ordered from SIS Förlag AB, who can also provide general information about Swedish and foreign standards.

SIS Förlag AB, SE 118 80 Stockholm, Sweden. Tel: +46 8 555 523 10. Fax: +46 8 555 523 11.

E-mail: [sis.sales@sis.se](mailto:sis.sales@sis.se) Internet: [www.sis.se](http://www.sis.se)



EUROPEAN STANDARD  
NORME EUROPÉENNE  
EUROPÄISCHE NORM

**EN 14133**

May 2009

ICS 67.160.10

Supersedes EN 14133:2003

English Version

## Foodstuffs - Determination of ochratoxin A in wine and beer - HPLC method with immunoaffinity column clean-up

Produits alimentaires - Dosage de l'ochratoxine A dans le  
vin et la bière - Méthode par purification sur colonne  
d'immuno-affinité suivie d'une analyse par chromatographie  
liquide haute performance (CLHP)

Lebensmittel - Bestimmung von Ochratoxin A in Wein und  
Bier - HPLC-Verfahren mit Reinigung an einer  
Immunoaffinitätssäule

This European Standard was approved by CEN on 24 May 2009.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**Management Centre: Avenue Marnix 17, B-1000 Brussels**

## Contents

Foreword.....	3
1 Scope .....	4
2 Normative references .....	4
3 Principle.....	4
4 Reagents.....	4
5 Apparatus .....	6
6 Procedure .....	7
7 HPLC analysis.....	7
8 Calculation.....	8
9 Precision.....	8
10 Test report .....	10
Annex A (informative) Precision data.....	11
Bibliography .....	13

## Foreword

This document (EN 14133:2009) has been prepared by Technical Committee CEN/TC 275 "Food analysis - Horizontal methods", the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by November 2009, and conflicting national standards shall be withdrawn at the latest by November 2009.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document will supersede EN 14133:2003.

The 2003 version has been updated with the inclusion of the corrigendum and some minor editorial improvements.

Annex A is informative.

**WARNING — Ochratoxin A is a potent nephrotoxin and liver toxin and has been reported to have immunosuppressant properties. It is classified by the International Agency for Research on Cancer (IARC) as possibly carcinogenic to humans (Group 2B). Acetonitrile is hazardous. Toluene is highly flammable and harmful. Observe appropriate safety precautions for handling such compounds. Gloves and safety glasses shall be worn at all times and all standard and sample preparation stages shall be carried out in a fume cupboard. Operation outside the fume cupboard, such as measurement of standards by UV spectrometer, shall be performed with the standard in closed containers. Decontamination procedures for laboratory wastes have been reported by the International Agency for Research on Cancer (IARC), see [1].**

According to the CEN/CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland and the United Kingdom.

## 1 Scope

This European Standard specifies a method for the determination of ochratoxin A content in wine and beer using immunoaffinity column clean up and high performance liquid chromatography (HPLC), see [2] and [3].

This method has been validated in an interlaboratory study according to AOAC International Guidelines [4] for collaborative study procedures to validate characteristics of a method of analysis for the determination of ochratoxin A in wine and beer via the analysis of naturally contaminated and spiked samples of wine and beer at levels ranging from 0,1 ng/ml to 3 ng/ml.

## 2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN ISO 3696, *Water for analytical laboratory use — Specification and test methods (ISO 3696:1987)*

## 3 Principle

Wine and beer samples are diluted with a solution containing polyethylene glycol (PEG) and sodium hydrogen carbonate, filtered and cleaned up by immunoaffinity column. Ochratoxin A is eluted with methanol and quantified by reversed-phase HPLC with fluorescence detection.

NOTE The use of PEG is essential to increase ochratoxin A recoveries and to reduce the number and intensity of other chromatographic peaks.

## 4 Reagents

### 4.1 General

Unless otherwise stated, use only reagents of recognized analytical grade and only distilled water or water of grade 1 as defined in EN ISO 3696.

Commercially available reagents with equivalent properties to the ones listed may be used.

### 4.2 Sodium chloride

### 4.3 Sodium hydrogen carbonate

### 4.4 Polyethylene glycol, (average molecular weight of 8000)

### 4.5 Methanol, HPLC grade

### 4.6 Acetonitrile, HPLC grade

### 4.7 Water, HPLC grade

### 4.8 Glacial acetic acid, $\varphi(\text{CH}_3\text{COOH}) \approx 99\%$

### 4.9 Diluting solution

Dissolve 10 g polyethylene glycol (4.4) and 50 g sodium hydrogen carbonate (4.3) in approximately 950 ml of water and bring up to 1000 ml with water.



#### 4.10 Washing solution

Dissolve 25 g sodium chloride (4.2) and 5 g sodium hydrogen carbonate (4.3) in approximately 950 ml of water and bring up to 1000 ml with water.

#### 4.11 HPLC mobile phase

Mix 990 ml HPLC water (4.7) with 990 ml acetonitrile (4.6) and 20 ml glacial acetic acid (4.8), filter through 0,45 µm filter and degas (e.g. with helium).

#### 4.12 Toluene

#### 4.13 Solvent mixture of toluene and glacial acetic acid

Mix 99 parts per volume of toluene (4.12) with 1 part per volume of glacial acetic acid (4.8).

#### 4.14 Ochratoxin A stock solution

Dissolve 1 mg of ochratoxin A (in crystal form) or the contents of 1 ampoule (if ochratoxin A has been obtained as a film) in solvent mixture (4.13) to give a solution containing approximately 20 µg/ml to 30 µg/ml of ochratoxin A. To determine the exact concentration, record the absorption curve between a wavelength of 300 nm and 370 nm in 5 nm steps in 1 cm quartz cells (5.10) and solvent mixture (4.13) as reference. Identify the wavelength for maximum absorption and calculate the mass concentration of ochratoxin A,  $\rho_{OTA}$ , in micrograms per millilitre, using Equation (1):

$$\rho_{ota} = \frac{A_{\max} \times M \times 100}{\epsilon \times b} \quad (1)$$

where

- $A_{\max}$  is the absorption determined at the maximum of the absorption curve (here: at 333 nm);
- $M$  is the molar mass of ochratoxin A ( $M = 403,8$  g/mol);
- $\epsilon$  is the molar absorption coefficient of ochratoxin A in the solvent mixture (4.13), (here: 544 m<sup>2</sup>/mol);
- $b$  is the optical path length of the quartz cell in centimetres.

This solution is stable at -18 °C for at least 4 years.

#### 4.15 Ochratoxin A standard solution

Dilute the stock solution (4.14) with solvent mixture (4.13) to obtain a standard solution with a mass concentration of ochratoxin A of 2 µg/ml. Store this solution in a refrigerator at approximately 4 °C and check the stability.

#### 4.16 Ochratoxin A calibration solution

Pipette 0,5 ml of the 2 µg/ml ochratoxin A standard solution (4.15) into a 10 ml volumetric flask (5.3) and evaporate the solvent under a stream of nitrogen. Redissolve in 10 ml of mobile phase (4.11). This gives a mass concentration of 100 ng/ml ochratoxin A.

Prepare six HPLC calibrants in separate 5 ml volumetric flasks (5.3) according to Table 1. Dilute each solution to 5 ml with HPLC mobile phase (4.11).

Table 1 — Preparation of calibration solutions

	Std 1	Std 2	Std 3	Std 4	Std 5	Std 6
µl of filtered mobile phase (4.11)	4970	4900	4700	4000	3000	2000
µl of 100 ng/ml OTA solution	30	100	300	1000	2000	3000
OTA mass concentration (ng/ml)	0,6	2,0	6,0	20	40	60
injected ng OTA	0,06	0,20	0,60	2,00	4,00	6,00

Prepare the calibration solutions at the beginning of every day of the analysis.

#### 4.17 Immunoaffinity columns

The immunoaffinity column shall contain antibodies raised against ochratoxin A. The column shall have a total capacity of not less than 100 ng of ochratoxin A and shall give a recovery of not less than 85 % when a diluted wine solution containing 100 ng of ochratoxin A is applied.

## 5 Apparatus

Usual laboratory equipment and, in particular, the following:

**5.1 Microbalance**, capable to measure 0,01 mg

**5.2 Glass vials**, approximately 4 ml, with polytetrafluoroethylene (PTFE)-lined screw cap, or appropriate sealable screw cap.

Certain types of vials can lead to losses of ochratoxin A during evaporation. To avoid this, silanization can be applied. Prepare vials by filling them with silanizing reagent and leave this reagent in the vial for 1 min. Rinse the vial twice with appropriate solvent (toluene, acetone or hexane) followed by water (twice) and dry the vial.

**5.3 Volumetric flasks**, 5 ml and 10 ml volume

**5.4 Vacuum manifold**, to accommodate immunoaffinity columns

**5.5 Reservoirs and attachments**, to fit to immunoaffinity columns

**5.6 Glass microfibre filters**, pore size 1,6 µm, (e.g. Whatman GF/A<sup>1</sup> or equivalent)

**5.7 Solvent evaporator**

**5.8 Calibrated microliter syringe(s) or microliter pipette(s)**

**5.9 HPLC apparatus consisting of**

---

<sup>1</sup> Whatman is an example of a suitable product available commercially. This information is given for the convenience of users of this European Standard and does not constitute an endorsement by CEN of these products. Equivalent products can be used if they can be shown to give equivalent results.

**5.9.1 Injection system**, a syringe-loading injection valve with 100 µl injection loop or equivalent.

**5.9.2 HPLC pump**, isocratic, capable of maintaining a volume flow rate of 1,0 ml/min.

**5.9.3 Analytical reverse phase separating column**, for example stainless steel (150 mm length, 4,6 mm inner diameter) packed with 5 µm C<sub>18</sub> reverse-phase material preceded by a suitable corresponding reverse-phase guard column or guard filter (0,5 µm). Columns of different dimensions can be used provided that they ensure a baseline resolution of the ochratoxin A peak from all other peaks.

**5.9.4 Fluorescence detector**, fitted with a flow cell and set at 333 nm (excitation) and 460 nm (emission). Detection of at least 0,02 ng of ochratoxin A shall be possible.

**5.9.5 Data system**

**5.10 UV spectrometer**, with suitable quartz cells

## 6 Procedure

### 6.1 Sample preparation

Degas sparkling wine and beer prior to dilution. Pour 10 ml of wine (beer) into a sealable 100 ml conical flask. Add 10 ml of the diluting solution (4.9) and mix vigorously. Filter through glass microfibre filter (5.6) when cloudy solutions or solid residues are formed after dilution.

If there is a problem with degassing, degassing should be performed by sonicating samples for 1 h, previously cooled at +4 °C for 30 min to prevent fast foam formation.

### 6.2 Immunoaffinity column clean-up

Prepare the immunoaffinity column according to the suppliers instructions.

Connect the immunoaffinity column (4.17) to the vacuum manifold (5.4) and attach the reservoir (5.5) to the immunoaffinity column. Add 10 ml (equivalent to 5 ml wine or beer) of the diluted solution to the reservoir and pass through the immunoaffinity column at a flow rate of about 1 drop per second. The immunoaffinity column shall not be allowed to run dry. Wash the immunoaffinity column with 5 ml of washing solution (4.10) and then with 5 ml of water at 1 to 2 drops per second flow rate. Dry the column by passing air through it. Remove the immunoaffinity column from the vacuum manifold and place it over a vial (5.2).

### 6.3 Preparation of the sample test solution

Elute ochratoxin A into the vial by passing 2 ml of methanol (4.5) at 1 drop per second flow rate. Evaporate the eluate to dryness under a stream of nitrogen e.g. at approximately 50 °C. Redissolve in 250 µl HPLC mobile phase (4.11) and store at 4 °C until HPLC analysis. This is the sample test solution.

## 7 HPLC analysis

### 7.1 Sample analysis

Inject 100 µl of reconstituted extract (equivalent to 2 ml wine or beer) into the chromatographic apparatus using the mobile phase (4.11) at flow rate of 1,0 ml/min.

### 7.2 Calibration curve

Prepare a calibration curve at the beginning of every day of the analysis and whenever chromatographic conditions change.