

SVENSK STANDARD

SS-EN ISO 16664:2018



Fastställt/Approved: 2018-01-22
Publicerad/Published: 2018-01-25
Utgåva/Edition: 2
Språk/Language: engelska/English
ICS: 71.040.40

Gasanalys – Hantering av kalibreringsgaser och gasblandningar – Riktlinjer (ISO 16664:2017)

Gas analysis – Handling of calibration gases and gas mixtures – Guidelines (ISO 16664:2017)

This preview is downloaded from www.sis.se. Buy the entire standard via <https://www.sis.se/std-80000679>

Standarder får världen att fungera

SIS (Swedish Standards Institute) är en fristående ideell förening med medlemmar från både privat och offentlig sektor. Vi är en del av det europeiska och globala nätverk som utarbetar internationella standarder. Standarder är dokumenterad kunskap utvecklad av framstående aktörer inom industri, näringsliv och samhälle och befrämjar handel över gränser, bidrar till att processer och produkter blir säkrare samt effektiviserar din verksamhet.

Delta och påverka

Som medlem i SIS har du möjlighet att påverka framtida standarder inom ditt område på nationell, europeisk och global nivå. Du får samtidigt tillgång till tidig information om utvecklingen inom din bransch.

Ta del av det färdiga arbetet

Vi erbjuder våra kunder allt som rör standarder och deras tillämpning. Hos oss kan du köpa alla publikationer du behöver – allt från enskilda standarder, tekniska rapporter och standardpaket till handböcker och onlinetjänster. Genom vår webbtjänst e-nav får du tillgång till ett lättnavigerat bibliotek där alla standarder som är aktuella för ditt företag finns tillgängliga. Standarder och handböcker är källor till kunskap. Vi säljer dem.

Utveckla din kompetens och lyckas bättre i ditt arbete

Hos SIS kan du gå öppna eller företagsinterna utbildningar kring innehåll och tillämpning av standarder. Genom vår närhet till den internationella utvecklingen och ISO får du rätt kunskap i rätt tid, direkt från källan. Med vår kunskap om standarders möjligheter hjälper vi våra kunder att skapa verklig nytta och lönsamhet i sina verksamheter.

Vill du veta mer om SIS eller hur standarder kan effektivisera din verksamhet är du välkommen in på www.sis.se eller ta kontakt med oss på tel 08-555 523 00.



Standards make the world go round

SIS (Swedish Standards Institute) is an independent non-profit organisation with members from both the private and public sectors. We are part of the European and global network that draws up international standards. Standards consist of documented knowledge developed by prominent actors within the industry, business world and society. They promote cross-border trade, they help to make processes and products safer and they streamline your organisation.

Take part and have influence

As a member of SIS you will have the possibility to participate in standardization activities on national, European and global level. The membership in SIS will give you the opportunity to influence future standards and gain access to early stage information about developments within your field.

Get to know the finished work

We offer our customers everything in connection with standards and their application. You can purchase all the publications you need from us - everything from individual standards, technical reports and standard packages through to manuals and online services. Our web service e-nav gives you access to an easy-to-navigate library where all standards that are relevant to your company are available. Standards and manuals are sources of knowledge. We sell them.

Increase understanding and improve perception

With SIS you can undergo either shared or in-house training in the content and application of standards. Thanks to our proximity to international development and ISO you receive the right knowledge at the right time, direct from the source. With our knowledge about the potential of standards, we assist our customers in creating tangible benefit and profitability in their organisations.

If you want to know more about SIS, or how standards can streamline your organisation, please visit www.sis.se or contact us on phone +46 (0)8-555 523 00



Europastandarden EN ISO 16664:2017 gäller som svensk standard. Detta dokument innehåller den officiella engelska versionen av EN ISO 16664:2017.

Denna standard ersätter SS-EN ISO 16664:2008, utgåva 1.

The European Standard EN ISO 16664:2017 has the status of a Swedish Standard. This document contains the official version of EN ISO 16664:2017.

This standard supersedes the Swedish Standard SS-EN ISO 16664:2008, edition 1.

© Copyright/Upphovsrätten till denna produkt tillhör SIS, Swedish Standards Institute, Stockholm, Sverige. Användningen av denna produkt regleras av slutanvändarlicensen som återfinns i denna produkt, se standardens sista sidor.

© Copyright SIS, Swedish Standards Institute, Stockholm, Sweden. All rights reserved. The use of this product is governed by the end-user licence for this product. You will find the licence in the end of this document.

Uppllysningar om sakinnehållet i standarden lämnas av SIS, Swedish Standards Institute, telefon 08-555 520 00. Standarder kan beställas hos SIS Förlag AB som även lämnar allmänna uppllysningar om svensk och utländsk standard.

Information about the content of the standard is available from the Swedish Standards Institute (SIS), telephone +46 8 555 520 00. Standards may be ordered from SIS Förlag AB, who can also provide general information about Swedish and foreign standards.

Denna standard är framtagen av kommittén för Gasanalys, SIS/TK 423/AG 02.

Har du synpunkter på innehållet i den här standarden, vill du delta i ett kommande revideringsarbete eller vara med och ta fram andra standarder inom området? Gå in på www.sis.se - där hittar du mer information.

EUROPEAN STANDARD

EN ISO 16664

NORME EUROPÉENNE

EUROPÄISCHE NORM

June 2017

ICS 71.040.40

Supersedes EN ISO 16664:2008

English Version

Gas analysis - Handling of calibration gases and gas mixtures - Guidelines (ISO 16664:2017)

Analyse des gaz - Mise en oeuvre des gaz et des mélanges de gaz pour étalonnage - Lignes directrices (ISO 16664:2017)

Gasanalyse - Handhabung von Kalibriergasen und Gasgemischen - Richtlinien (ISO 16664:2017)

This European Standard was approved by CEN on 19 September 2016.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

SS-EN ISO 16664:2018 (E)

Contents		Page
European foreword		iv
Introduction		v
1	Scope	1
2	Normative references	1
3	Terms and definitions	1
4	Transport and storage	3
4.1	General remarks	3
4.2	Low temperature	3
4.3	High temperature	4
4.4	Water	4
4.5	Storage and handling	4
5	Mode of withdrawal	4
5.1	General	4
5.2	Minimum utilization pressure	4
5.3	Temperature	5
5.4	Pressure reduction and flow	5
5.5	Replacement, change of cylinder positions	5
6	Transfer system	5
6.1	Purging procedure	5
6.2	Considerations when designing and constructing gas transfer lines	6
6.2.1	Modes of gas sampling	6
6.2.2	Pressure- and flow-reducing equipment	7
6.2.3	Material for the construction of transfer lines	7
6.2.4	General methods and examples of sampling systems	10
7	Stability	13
Annex A (informative) Check on the stability of calibration gas mixtures by end-users		14
Bibliography		16

European foreword

This document (EN ISO 16664:2017) has been prepared by Technical Committee ISO/TC 158 "Analysis of gases".

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by December 2017, and conflicting national standards shall be withdrawn at the latest by December 2017.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 16664:2008.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 16664:2017 has been approved by CEN as EN ISO 16664:2017 without any modification.

SS-EN ISO 16664:2018 (E)

Introduction

This document uses the terms “calibration gas” for both gas mixtures and pure gases as the limiting case of gas mixtures.

The quality of calibration gases in cylinders as certified by producers is defined by

- a) the correct analyte content;
- b) a known uncertainty which is appropriate for its intended use;
- c) the stability;
- d) the homogeneity.

During its utilization period, the quality of calibration gases is influenced by

- storage conditions at the manufacturer’s and user’s sites;
- transport conditions;
- modes of calibration gas withdrawal and transfer;
- the transfer system employed.

Gas analysis — Handling of calibration gases and gas mixtures — Guidelines

SAFETY PRECAUTIONS — National and international safety regulations concerning storage, use and transportation of pure gases and gas mixtures are to be followed in addition to this document.

1 Scope

This document describes factors that may influence the composition of pure gases and homogeneous gas mixtures used for calibration purposes. This document only applies to gases or gas mixtures that are within the “utilization period”. It provides the following guidelines for the handling and use of calibration gas mixtures:

- storage of calibration gas cylinders;
- calibration gas withdrawal from cylinders;
- transfer of calibration gas from cylinders to the point of calibration.

It also outlines a method of assessing the stability of a gas mixture, taking into account the gas composition uncertainty given on the certificate and the user’s measurement uncertainty.

2 Normative references

There are no normative references in this document.

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

3.1

calibration gas

pure gas or gas mixture used for calibration

3.2

calibration gas mixture

gas mixture of known *stability* (3.9) and *homogeneity* (3.4) whose composition is well established for use in the calibration or verification of a measuring instrument or for the validation of a measurement

Note 1 to entry: Calibration gas mixtures are measurement standards ([Annex A](#)) as defined in ISO/IEC Guide 99:2007.

[SOURCE: ISO 7504:2015, 5.1]

3.3

component

chemical entity at a defined physical state present in a material or in a mixture

[SOURCE: ISO 7504:2015, 3.3]

SS-EN ISO 16664:2018 (E)

3.4

homogeneity

state of a gas mixture wherein all of its components are distributed uniformly throughout the volume occupied by the gas mixture

[SOURCE: ISO 7504:2015, 3.1]

3.5

impurity

undesired minor component present in a parent gas (and thus detectable in a gas mixture made of this parent gas)

[SOURCE: ISO 7504:2015, 5.5]

3.6

leak rate

volume of fluid leaking from the system per unit of time due to incomplete sealing of materials

3.7

leak tightness

conformity to a specified leak rate

3.8

step response time

duration between the instant when an input quantity value of a measuring instrument or measuring system is subjected to an abrupt change between two specified constant quantity values and the instant when a corresponding indication settles within specified limits around its final steady value

[SOURCE: ISO/IEC Guide 99:2007]

3.9

stability

attribute of a gas mixture, under specified conditions, to maintain its composition within specified uncertainty limits for a specified period of time (maximum storage life)

[SOURCE: ISO 7504:2015, 3.2]

3.10

maximum storage life

period after which the properties stated for a gas mixture cannot be warranted to lie within their limits

Note 1 to entry: This period is usually identified as that for which the producer ensures that the gas mixture maintains its composition within the specified limits when it is stored in accordance with the requirements based upon the concepts defined in ISO 7504:2015, 7.1 to 7.4.

Note 2 to entry: The end of this period may be indicated by an "expiry date" (see also ISO 6142-1).

[SOURCE: ISO 7504:2015, 7.5]

3.11

transfer system

gas-conducting system which begins at the cylinder valve and ends at the gas sample inlet to the measuring instrument and includes all structural elements

3.12

measurement uncertainty

non-negative parameter characterizing the dispersion of the quantity values being attributed to a measurand, based on the information used

Note 1 to entry: Measurement uncertainty includes components arising from systematic effects, such as components associated with corrections and the assigned quantity values of measurement standards, as well as the definitional uncertainty. Sometimes estimated systematic effects are not corrected for but, instead, associated measurement uncertainty components are incorporated.

Note 2 to entry: The parameter may be, for example, a standard deviation called standard measurement uncertainty (or a specified multiple of it), or the half-width of an interval, having a stated coverage probability.

Note 3 to entry: Measurement uncertainty comprises, in general, many components. Some of these may be evaluated by Type A evaluation of measurement uncertainty from the statistical distribution of the quantity values from series of measurements and can be characterized by standard deviations. The other components, which may be evaluated by Type B evaluation of measurement uncertainty, can also be characterized by standard deviations, evaluated from probability density functions based on experience or other information.

Note 4 to entry: In general, for a given set of information, it is understood that the measurement uncertainty is associated with a stated quantity value attributed to the measurand. A modification of this value results in a modification of the associated uncertainty.

[SOURCE: ISO/IEC Guide 99:2007, 2.26]

3.13

utilization period

time between the date of certification and the expiry date

3.14

permeability

property of a material of transmitting gases and liquids by passage through one surface and out at another surface by diffusion and sorption processes

Note 1 to entry: Not to be confused with porosity.

[SOURCE: ISO 472:2013, 2.690]

4 Transport and storage

4.1 General remarks

After preparation of the calibration gas, the gas cylinder is transported to the user. The environment in which the cylinders are transported is not normally regulated in terms of temperature and humidity. Low temperatures may have a detrimental impact on the mixture composition, especially when condensable components are present in the mixture. As a consequence of this, environmental conditions during transport and storage should never exceed those recommended by the manufacturer.

Gas cylinders and especially cylinder valves shall be free of grease and other lubricants. During storage and transportation, cylinder valves shall be closed, sealing nuts or plugs shall be tightened and suitable protection devices shall be attached, e.g. caps or valve guards.

The gas cylinders are transported in several ways, e.g. by air, railway, road and on water. In some specific cases, the temperature restrictions may be such that not every means of transport is acceptable.

4.2 Low temperature

The gas cylinder may be exposed to low temperatures during storage and transportation. For gas mixtures containing condensable components, it is important that the cylinder is not stored or transported at temperatures below those recommended by the manufacturer. If the mixture is exposed