

# SVENSK STANDARD

## SS-EN 868-5:2019



Fastställt/Approved: 2019-01-07  
Utgåva/Edition: 3  
Språk/Language: engelska/English  
ICS: 11.080.30;55.040

---

### **Förpackningsmaterial för medicintekniska produkter avsedda för sterilisering i sluten förpackning – Del 5: Förseglingsbara påsar och rullar av porösa material och plastfilm – Krav och provningsmetoder**

### **Packaging for terminally sterilized medical devices – Part 5: Sealable pouches and reels of porous materials and plastic film construction – Requirements and test methods**

This preview is downloaded from [www.sis.se](http://www.sis.se). Buy the entire standard via <https://www.sis.se/std-80009249>

# Standarder får världen att fungera

*SIS (Swedish Standards Institute) är en fristående ideell förening med medlemmar från både privat och offentlig sektor. Vi är en del av det europeiska och globala nätverk som utarbetar internationella standarder. Standarder är dokumenterad kunskap utvecklad av framstående aktörer inom industri, näringsliv och samhälle och befrämjar handel över gränser, bidrar till att processer och produkter blir säkrare samt effektiviserar din verksamhet.*

## Delta och påverka

Som medlem i SIS har du möjlighet att påverka framtida standarder inom ditt område på nationell, europeisk och global nivå. Du får samtidigt tillgång till tidig information om utvecklingen inom din bransch.

## Ta del av det färdiga arbetet

Vi erbjuder våra kunder allt som rör standarder och deras tillämpning. Hos oss kan du köpa alla publikationer du behöver – allt från enskilda standarder, tekniska rapporter och standardpaket till handböcker och onlinetjänster. Genom vår webbtjänst e-nav får du tillgång till ett lättnavigerat bibliotek där alla standarder som är aktuella för ditt företag finns tillgängliga. Standarder och handböcker är källor till kunskap. Vi säljer dem.

## Utveckla din kompetens och lyckas bättre i ditt arbete

Hos SIS kan du gå öppna eller företagsinterna utbildningar kring innehåll och tillämpning av standarder. Genom vår närhet till den internationella utvecklingen och ISO får du rätt kunskap i rätt tid, direkt från källan. Med vår kunskap om standarders möjligheter hjälper vi våra kunder att skapa verklig nytta och lönsamhet i sina verksamheter.

**Vill du veta mer om SIS eller hur standarder kan effektivisera din verksamhet är du välkommen in på [www.sis.se](http://www.sis.se) eller ta kontakt med oss på tel 08-555 523 00.**



# Standards make the world go round

*SIS (Swedish Standards Institute) is an independent non-profit organisation with members from both the private and public sectors. We are part of the European and global network that draws up international standards. Standards consist of documented knowledge developed by prominent actors within the industry, business world and society. They promote cross-border trade, they help to make processes and products safer and they streamline your organisation.*

## Take part and have influence

As a member of SIS you will have the possibility to participate in standardization activities on national, European and global level. The membership in SIS will give you the opportunity to influence future standards and gain access to early stage information about developments within your field.

## Get to know the finished work

We offer our customers everything in connection with standards and their application. You can purchase all the publications you need from us - everything from individual standards, technical reports and standard packages through to manuals and online services. Our web service e-nav gives you access to an easy-to-navigate library where all standards that are relevant to your company are available. Standards and manuals are sources of knowledge. We sell them.

## Increase understanding and improve perception

With SIS you can undergo either shared or in-house training in the content and application of standards. Thanks to our proximity to international development and ISO you receive the right knowledge at the right time, direct from the source. With our knowledge about the potential of standards, we assist our customers in creating tangible benefit and profitability in their organisations.

**If you want to know more about SIS, or how standards can streamline your organisation, please visit [www.sis.se](http://www.sis.se) or contact us on phone +46 (0)8-555 523 00**



Europastandarden EN 868-5:2018 gäller som svensk standard. Detta dokument innehåller den officiella engelska versionen av EN 868-5:2018.

Denna standard ersätter SS-EN 868-5:2009, utgåva 2.

The European Standard EN 868-5:2018 has the status of a Swedish Standard. This document contains the official version of EN 868-5:2018.

This standard supersedes the SS-EN 868-5:2009, edition 2.

© Copyright/Upphovsrätten till denna produkt tillhör SIS, Swedish Standards Institute, Stockholm, Sverige. Användningen av denna produkt regleras av slutanvändarlicensen som återfinns i denna produkt, se standardens sista sidor.

© Copyright SIS, Swedish Standards Institute, Stockholm, Sweden. All rights reserved. The use of this product is governed by the end-user licence for this product. You will find the licence in the end of this document.

*Upplysningar om sakinnehållet i standarden lämnas av SIS, Swedish Standards Institute, telefon 08-555 520 00. Standarder kan beställas hos SIS som även lämnar allmänna upplysningar om svensk och utländsk standard.*

*Information about the content of the standard is available from the Swedish Standards Institute (SIS), telephone +46 8 555 520 00. Standards may be ordered from SIS, who can also provide general information about Swedish and foreign standards.*

Denna standard är framtagen av kommittén för Rengöring, desinfektion och sterilisering, SIS/TK 349.

Har du synpunkter på innehållet i den här standarden, vill du delta i ett kommande revideringsarbete eller vara med och ta fram andra standarder inom området? Gå in på [www.sis.se](https://www.sis.se) - där hittar du mer information.



EUROPEAN STANDARD

**EN 868-5**

NORME EUROPÉENNE

EUROPÄISCHE NORM

December 2018

ICS 11.080.30

Supersedes EN 868-5:2009

English Version

## Packaging for terminally sterilized medical devices - Part 5: Sealable pouches and reels of porous materials and plastic film construction - Requirements and test methods

Emballages des dispositifs médicaux stérilisés au stade terminal - Partie 5: Sachets et gaines scellables constitués d'une face matière poreuse et d'une face film plastique - Exigences et méthodes d'essai

Verpackungen für in der Endverpackung zu sterilisierende Medizinprodukte - Teil 5: Siegelfähige Klarsichtbeutel und -schläuche aus porösen Materialien und Kunststoff-Verbundfolie - Anforderungen und Prüfverfahren

This European Standard was approved by CEN on 20 August 2018.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Rue de la Science 23, B-1040 Brussels**

**SS-EN 868-5:2019 (E)**

<b>Contents</b>		Page
<b>European foreword</b> .....		3
<b>Introduction</b> .....		5
<b>1</b>	<b>Scope</b> .....	6
<b>2</b>	<b>Normative references</b> .....	6
<b>3</b>	<b>Terms and definitions</b> .....	6
<b>4</b>	<b>Requirements</b> .....	7
<b>4.1</b>	<b>General</b> .....	7
<b>4.2</b>	<b>Materials</b> .....	7
<b>4.2.1</b>	<b>Porous material</b> .....	7
<b>4.2.2</b>	<b>Plastic film</b> .....	7
<b>4.3</b>	<b>Construction and design</b> .....	8
<b>4.4</b>	<b>Process indicator</b> .....	8
<b>4.5</b>	<b>Performance requirements and test methods</b> .....	8
<b>4.6</b>	<b>Marking</b> .....	9
<b>4.6.1</b>	<b>Pouches and reels</b> .....	9
<b>4.6.2</b>	<b>Transport and/or storage packaging</b> .....	10
<b>5</b>	<b>Information to be supplied by the manufacturer</b> .....	10
<b>Annex A (informative) Details of significant technical changes between this document and the previous edition</b> .....		11
<b>Annex B (normative) Method for the determination of resistance to the intended sterilization process</b> .....		12
<b>Annex C (normative) Method for the determination of pinholes in plastic laminate</b> .....		13
<b>Annex D (normative) Method for the determination of the strength of the seal for pouches and reel material</b> .....		15
<b>Annex E (normative) Method for the determination of peel characteristics of paper/plastic laminate products</b> .....		18
<b>Annex F (normative) Method for the determination of fibre orientation</b> .....		20
<b>Bibliography</b> .....		21

## European foreword

This document (EN 868-5:2018) has been prepared by Technical Committee CEN/TC 102 “Sterilizers and associated equipment for processing of medical devices”, the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by June 2019, and conflicting national standards shall be withdrawn at the latest by June 2019.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN 868-5:2009.

Annex A provides details of significant technical changes between this European Standard and the previous edition.

EN 868 consists of the following parts, under the general title *Packaging for terminally sterilized medical devices*:

- *Part 2: Sterilization wrap — Requirements and test methods;*
- *Part 3: Paper for use in the manufacture of paper bags (specified in EN 868-4) and in the manufacture of pouches and reels (specified in EN 868-5) — Requirements and test methods;*
- *Part 4: Paper bags — Requirements and test methods;*
- *Part 5: Sealable pouches and reels of porous materials and plastic film construction — Requirements and test methods;*
- *Part 6: Paper for low temperature sterilization processes — Requirements and test methods;*
- *Part 7: Adhesive coated paper for low temperature sterilization processes — Requirements and test methods;*
- *Part 8: Re-usable sterilization containers for steam sterilizers conforming to EN 285 — Requirements and test methods;*
- *Part 9: Uncoated nonwoven materials of polyolefines — Requirements and test methods;*
- *Part 10: Adhesive coated nonwoven materials of polyolefines — Requirements and test methods.*

In addition, ISO/TC 198 “Sterilization of health care products” in collaboration with CEN/TC 102 “Sterilizers and associated equipment for processing of medical devices” has prepared the EN ISO 11607 series “Packaging for terminally sterilized medical devices”. The EN ISO 11607 series specifies general requirements for materials, sterile barrier systems and packaging systems (Part 1) and validation requirements for forming, sealing and assembly processes (Part 2).

According to the CEN-CENELEC Internal Regulations, the national standards organisations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta,

**SS-EN 868-5:2019 (E)**

Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.



## **Introduction**

The EN ISO 11607 series consists of two parts under the general title “Packaging for terminally sterilized medical devices”. Part 1 of this series specifies general requirements and test methods for materials, preformed sterile barrier systems, sterile barrier systems and packaging systems that are intended to maintain sterility of terminally sterilized medical devices to the point of use. Part 2 of this series specifies validation requirements for forming, sealing and assembly processes.

General requirements for all types of sterile barrier systems are provided by EN ISO 11607-1.

The EN 868 series can be used to demonstrate compliance with one or more of the requirements specified in EN ISO 11607-1.

CEN/TC 102/WG 4 also appreciates the initiatives of CEN with regard to the minimization of adverse environmental impacts by standards. It was agreed that this subject should be given priority during the next edition of the EN ISO 11607 series that is the basic reference for all parts of the EN 868 series.

## SS-EN 868-5:2019 (E)

### 1 Scope

This document specifies test methods and values for sealable pouches and reels manufactured from porous materials complying with either EN 868 part 2, 3, 6, 7, 9 or 10 and plastic film complying with Clause 4. These sealable pouches and reels are used as sterile barrier systems and/or packaging systems that are intended to maintain sterility of terminally sterilized medical devices to the point of use.

Other than the general requirements as specified in EN ISO 11607-1 and EN ISO 11607-2 this part of EN 868 specifies materials, test methods and values that are specific to the products covered by this document.

The materials specified in this part of EN 868 are intended for single use only.

### 2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 868-2:2017, *Packaging for terminally sterilized medical devices — Part 2: Sterilization wrap — Requirements and test methods*

EN 868-3:2017, *Packaging for terminally sterilized medical devices — Part 3: Paper for use in the manufacture of paper bags (specified in EN 868-4) and in the manufacture of pouches and reels (specified in EN 868-5) — Requirements and test methods*

EN 868-6:2017, *Packaging for terminally sterilized medical devices — Part 6: Paper for low temperature sterilization processes — Requirements and test methods*

EN 868-7:2017, *Packaging for terminally sterilized medical devices — Part 7: Adhesive coated paper for low temperature sterilization processes — Requirements and test methods*

EN 868-9:2018, *Packaging for terminally sterilized medical devices — Part 9: Uncoated nonwoven materials of polyolefines — Requirements and test methods*

EN 868-10:2018, *Packaging for terminally sterilized medical devices — Part 10: Adhesive coated nonwoven materials of polyolefines — Requirements and test methods*

EN ISO 11140-1, *Sterilization of health care products — Chemical indicators — Part 1: General requirements (ISO 11140-1)*

EN ISO 11607-1:2017, *Packaging for terminally sterilized medical devices — Part 1: Requirements for materials, sterile barrier systems and packaging systems (ISO 11607-1:2006, including Amd 1:2014)*

ISO 8601, *Data elements and interchange formats — Information interchange — Representation of dates and times*

ASTM D882:2012, *Test Methods for Tensile Properties of the Thin Plastic Sheeting*

ASTM F88/F88M:2015, *Standard Test Method for Seal Strength of Flexible Barrier Materials*

### 3 Terms and definitions

For the purposes of this document, the terms and definitions given in EN ISO 11607-1:2017 apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <http://www.iso.org/obp>

## **4 Requirements**

### **4.1 General**

For any material, preformed sterile barrier system or sterile barrier system, the requirements of EN ISO 11607-1 shall apply.

This part of EN 868 only introduces performance requirements and test methods that are specific to the products covered by this part of EN 868 but does not add or modify the general requirements specified in EN ISO 11607-1.

As such, the particular requirements in 4.5 can be used to demonstrate compliance with one or more but not all of the requirements of EN ISO 11607-1.

NOTE 1 Compliance to EN 868-5 does not automatically mean compliance to EN ISO 11607-1.

A confirmation of compliance to EN 868-5 shall contain a statement whether EN ISO 11607-1 is covered.

NOTE 2 When additional materials are used inside the sterile barrier system in order to ease the organization, drying or aseptic presentation (e.g. inner wrap, indicators, packing lists, mats, instrument organizer sets, tray liners or an additional envelope around the medical device) then other requirements, including the determination of the acceptability of these materials during validation activities, can apply.

### **4.2 Materials**

#### **4.2.1 Porous material**

The porous material shall comply with the requirements of Clause 4 of either EN 868-2:2017, EN 868-3:2017, EN 868-6:2017, EN 868-7:2017, EN 868-9:2018 or EN 868-10:2018.

If the intended method of sterilization is irradiation only, the requirements for wet strength properties and permeability to air for porous materials are not applicable.

#### **4.2.2 Plastic film**

**4.2.2.1** The plastic film shall be a composite of two or more layers. When tested after the intended sterilization process in accordance with Annex B the plastics interply bond shall not separate nor become cloudy.

**4.2.2.2** The plastic film shall be free from pinholes when tested in accordance with Annex C.

**4.2.2.3** When examined by unaided normal or corrected vision in transmitted light (daylight or good artificial light) the plastic film shall be free from foreign matter and/or other imperfections that would adversely affect compliance with the requirements of 4.5.

NOTE Slight continuous surface irregularities arising from the extrusion of the plastic film is not regarded as a defect.

**4.2.2.4** The plastic film shall be sealable to the porous material under the conditions specified.