

SVENSK STANDARD

SS-EN ISO 7886-3:2020

**Sterila injektionssprutor för engångsbruk –
Del 3: Självförstörande sprutor med bestämd dos, avsedda för
immunisering (ISO 7886-3:2020)**

**Sterile hypodermic syringes for single use –
Part 3: Auto-disabled syringes for fixed-dose immunization
(ISO 7886-3:2020)**



sis Svenska
Institutet för
Standarder

Language: engelska/English

Edition: 3

This preview is downloaded from www.sis.se. Buy the entire standard via <https://www.sis.se/std-80022148>

Den här standarden kan hjälpa dig att effektivisera och kvalitetssäkra ditt arbete. SIS har fler tjänster att erbjuda dig för att underlätta tillämpningen av standarder i din verksamhet.

SIS Abonnemang

Snabb och enkel åtkomst till gällande standard med SIS Abonnemang, en prenumerationstjänst genom vilken din organisation får tillgång till all världens standarder, senaste uppdateringarna och där hela din organisation kan ta del av innehållet i prenumerationen.

Utbildning, event och publikationer

Vi erbjuder även utbildningar, rådgivning och event kring våra mest sålda standarder och frågor kopplade till utveckling av standarder. Vi ger också ut handböcker som underlättar ditt arbete med att använda en specifik standard.

Vill du delta i ett standardiseringsprojekt?

Genom att delta som expert i någon av SIS 300 tekniska kommittéer inom CEN (europeisk standardisering) och/eller ISO (internationell standardisering) har du möjlighet att påverka standardiseringsarbetet i frågor som är viktiga för din organisation. Välkommen att kontakta SIS för att få veta mer!

Kontakt

Skriv till kundservice@sis.se, besök [sis.se](https://www.sis.se) eller ring 08 - 555 523 10

© Copyright/Upphovsrätten till denna produkt tillhör Svenska institutet för standarder, Stockholm, Sverige. Upphovsrätten och användningen av denna produkt regleras i slutanvändarlicensen som återfinns på [sis.se/slutanvandarlicens](https://www.sis.se/slutanvandarlicens) och som du automatiskt blir bunden av när du använder produkten. För ordlista och förkortningar se [sis.se/ordlista](https://www.sis.se/ordlista).

© Copyright Svenska institutet för standarder, Stockholm, Sweden. All rights reserved. The copyright and use of this product is governed by the end-user licence agreement which you automatically will be bound to when using the product. You will find the licence at [sis.se/enduserlicenseagreement](https://www.sis.se/enduserlicenseagreement).

Upplysningar om sakinnehållet i standarden lämnas av Svenska institutet för standarder, telefon 08 - 555 520 00. Standarder kan beställas hos SIS som även lämnar allmänna upplysningar om svensk och utländsk standard.

Standarden är framtagen av kommittén för Förbrukningsmaterial inom sjukvården, SIS/TK 330.

Har du synpunkter på innehållet i den här standarden, vill du delta i ett kommande revideringsarbete eller vara med och ta fram andra standarder inom området? Gå in på www.sis.se - där hittar du mer information.

Europastandarden EN ISO 7886-3:2020 gäller som svensk standard. Detta dokument innehåller den officiella engelska versionen av EN ISO 7886-3:2020.

Denna standard ersätter SS-EN ISO 7886-3:2009, utgåva 2

The European Standard EN ISO 7886-3:2020 has the status of a Swedish Standard. This document contains the official version of EN ISO 7886-3:2020.

This standard supersedes the SS-EN ISO 7886-3:2009, edition 2

EUROPEAN STANDARD

EN ISO 7886-3

NORME EUROPÉENNE

EUROPÄISCHE NORM

May 2020

ICS 11.040.25

Supersedes EN ISO 7886-3:2009

English Version

**Sterile hypodermic syringes for single use - Part 3:
Auto-disabled syringes for fixed-dose immunization
(ISO 7886-3:2020)**

Seringues hypodermiques stériles, non réutilisables
- Partie 3: Seringues autobloquantes pour
vaccination à dose fixe (ISO 7886-3:2020)

Sterile Einmalspritzen für medizinische Zwecke -
Teil 3: Selbstblockierende Spritzen für die Injektion
mit fixer Impfstoffdosis (ISO 7886-3:2020)

This European Standard was approved by CEN on 24 April 2020.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents

Page

Foreword	viii
European foreword	ix
Introduction	x
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Nomenclature	2
5 General requirements	3
6 Extraneous matter	4
6.1 General	4
6.2 Limits for acidity or alkalinity	4
6.3 Limits for extractable metals	4
7 Lubricant	4
8 Tolerance on nominal capacity	4
9 Graduated scale	5
9.1 Scale.....	5
9.2 Position of scale.....	5
10 Barrel	5
10.1 Dimensions.....	5
10.2 Barrel flanges.....	5
11 Plunger stopper/plunger assembly	5
11.1 Design	5
11.2 Fit of the plunger stopper/plunger in the barrel.....	6
11.3 Fiducial line	6
12 Needle	6
12.1 General	6
12.2 Integrated needle.....	6
12.3 Non-integrated needle.....	6
12.4 Sharps protection features	6
13 Performance	7
13.1 General	7
13.2 Dead space.....	7
13.3 Freedom from air and liquid leakage	7
13.4 Auto-disable syringe feature.....	7
13.5 Performance after shipping.....	7
14 Packaging	8
14.1 Unit packaging providing sterile barrier	8
14.2 Multiple unit pack.....	8
14.3 User packaging.....	8
15 Information supplied by the manufacturer	8
15.1 General	8
15.2 Syringes.....	8
15.3 Unit packaging providing sterile barrier	8
15.4 User packaging.....	9
15.5 Storage containers.....	9
15.6 Transport wrapping	10
Annex A (normative) Method for preparation of extracts	11

Annex B (informative) Test method for forces required to operate piston	12
Annex C (normative) Test method for testing auto-disable syringe feature	14
Bibliography	15

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular, the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation of the voluntary nature of standards, the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the World Trade Organization (WTO) principles in the Technical Barriers to Trade (TBT), see www.iso.org/iso/foreword.html.

This document was prepared by Technical Committee ISO/TC 84, *Devices for administration of medicinal products and catheters*, in collaboration with the European Committee for Standardization (CEN) Technical Committee CEN/TC 205, *Non-active medical devices*, in accordance with the Agreement on technical cooperation between ISO and CEN (Vienna Agreement).

This second edition cancels and replaces the first edition (ISO 7886-3:2005), which has been technically revised. The main changes compared to the previous edition are as follows:

- update of the references, mainly ISO 7886-1:2017.

A list of all parts in the ISO 7886 series can be found on the ISO website.

Any feedback or questions on this document should be directed to the user's national standards body. A complete listing of these bodies can be found at www.iso.org/members.html.

European foreword

This document (EN ISO 7886-3:2020) has been prepared by Technical Committee ISO/TC 84 "Devices for administration of medicinal products and catheters" in collaboration with Technical Committee CEN/TC 205 "Non-active medical devices" the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by November 2020, and conflicting national standards shall be withdrawn at the latest by November 2020.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 7886-3:2009.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Republic of North Macedonia, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 7886-3:2020 has been approved by CEN as EN ISO 7886-3:2020 without any modification.

Introduction

ISO 7886 was first published in 1984. It was subsequently decided to divide it into two parts: ISO 7886-1 retaining essentially the scope of ISO 7886:1984 and ISO 7886-2 being applicable to sterile, single-use syringes for use with power-driven pumps.

The preparation of this document was recognized as a high priority to prevent the reuse of fixed dose immunization syringes. Reuse of injection equipment in the absence of sterilization has increasingly led to transmission of blood-borne pathogens.

The World Health Organization (WHO) had produced a specification for syringes that are rendered inactive after one use (commonly referred to as “auto-disabled” syringes). It was agreed that an additional part of the ISO 7886 series would be needed to cover “auto-disabled” syringes, while leaving in place ISO 7886-1 and ISO 7886-2 without modification, as a large number of devices in common use would not be intended to conform with the auto-disable properties suggested.

It has been discussed to limit the syringe types to only comprise the type having an auto-disable syringe feature that is automatically activated and remains effective from the time that the injection is commenced. An assessment of potential hazards based only on hypothetical use indicates that the type having an auto-disable syringe feature that is automatically activated and remains effective from the time of the injection being initiated is potentially safer than the other types. However, no consensus could be reached on either deleting types or retaining them, as no reliable risk data from field use exists at present. It was therefore agreed to retain all types and restrict this revision to alignment with ISO 7886-1:2017 and initiate a new revision if new field studies or incident reports indicate a need for a revision.

It is recognized that syringes designed to reduce the risk of needle stick injuries can also conform with this document.

In some countries national regulations might take precedence over the requirements in this document.

Guidance on transition periods for implementing the requirements of this document is given in ISO/TR 19244.

Sterile hypodermic syringes for single use —

Part 3: Auto-disabled syringes for fixed-dose immunization

1 Scope

This document specifies the properties and performance of sterile single-use hypodermic syringes with an auto-disable syringe feature intended to deliver a fixed dose of vaccine immediately after filling. The syringes can be made of plastic, rubber or other materials and can be with or without needle and needle protection feature.

This document does not specify the design of the auto-disable syringe feature.

This document is not applicable to syringes for use with insulin (covered by ISO 8537), syringes for use with power-driven syringe pumps (covered by ISO 7886-2), reuse prevention syringes (covered by ISO 7886-4) or syringes designed to be prefilled (covered by the ISO 11040 series). It does not address compatibility with injection fluids/vaccines.

2 Normative references

The following documents are referred to in the text in such a way that some or all of their content constitutes requirements of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 7864:2016, *Sterile hypodermic needles for single use — Requirements and test methods*

ISO 7886-1:2017, *Sterile hypodermic syringes for single use — Part 1: Syringes for manual use*

ISO 8537:2016, *Sterile single-use syringes, with or without needle, for insulin*

ISO 9626, *Stainless steel needle tubing for the manufacture of medical devices — Requirements and test methods*

ISO 10993-1, *Biological evaluation of medical devices — Part 1: Evaluation and testing within a risk management process*

ISO 14971, *Medical devices — Application of risk management to medical devices*

ISO 23908, *Sharps injury protection — Requirements and test methods — Sharps protection features for single-use hypodermic needles, introducers for catheters and needles used for blood sampling*

ISO 80369-7, *Small-bore connectors for liquids and gases in healthcare applications — Part 7: Connectors for intravascular or hypodermic applications*

ASTM D999-01, *Standard methods for vibration testing of shipping containers*

ASTM D5276-98, *Standard test method for drop test of loaded containers by free fall*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in ISO 7886-1 and ISO 8537 and the following apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

— ISO Online browsing platform: available at <https://www.iso.org/obp>