

SVENSK STANDARD

SS-EN ISO 14253-5:2015



Fastställt/Approved: 2015-09-21
Publicerad/Published: 2015-10-06
Utgåva/Edition: 1
Språk/Language: engelska/English
ICS: 17.040.01

Geometrisk produktspecifikation (GPS) – Kontroll genom mätning av arbetsstycken och mätinstrument – Del 5: Osäkerhet i test indikerar mätning (ISO 14253-5:2015)

Geometrical product specifications (GPS) – Inspection by measurement of workpieces and measuring equipment – Part 5: Uncertainty in verification testing of indicating measuring instruments (ISO 14253-5:2015)

This preview is downloaded from www.sis.se. Buy the entire standard via <https://www.sis.se/std-8016285>

Standarder får världen att fungera

SIS (Swedish Standards Institute) är en fristående ideell förening med medlemmar från både privat och offentlig sektor. Vi är en del av det europeiska och globala nätverk som utarbetar internationella standarder. Standarder är dokumenterad kunskap utvecklad av framstående aktörer inom industri, näringsliv och samhälle och befrämjar handel över gränser, bidrar till att processer och produkter blir säkrare samt effektiviserar din verksamhet.

Delta och påverka

Som medlem i SIS har du möjlighet att påverka framtida standarder inom ditt område på nationell, europeisk och global nivå. Du får samtidigt tillgång till tidig information om utvecklingen inom din bransch.


Ta del av det färdiga arbetet

Vi erbjuder våra kunder allt som rör standarder och deras tillämpning. Hos oss kan du köpa alla publikationer du behöver – allt från enskilda standarder, tekniska rapporter och standardpaket till handböcker och onlinetjänster. Genom vår webbtjänst e-nav får du tillgång till ett lättnavigerat bibliotek där alla standarder som är aktuella för ditt företag finns tillgängliga. Standarder och handböcker är källor till kunskap. Vi säljer dem.

Utveckla din kompetens och lyckas bättre i ditt arbete

Hos SIS kan du gå öppna eller företagsinterna utbildningar kring innehåll och tillämpning av standarder. Genom vår närhet till den internationella utvecklingen och ISO får du rätt kunskap i rätt tid, direkt från källan. Med vår kunskap om standarders möjligheter hjälper vi våra kunder att skapa verklig nytta och lönsamhet i sina verksamheter.

Vill du veta mer om SIS eller hur standarder kan effektivisera din verksamhet är du välkommen in på www.sis.se eller ta kontakt med oss på tel 08-555 523 00.



Standards make the world go round

SIS (Swedish Standards Institute) is an independent non-profit organisation with members from both the private and public sectors. We are part of the European and global network that draws up international standards. Standards consist of documented knowledge developed by prominent actors within the industry, business world and society. They promote cross-border trade, they help to make processes and products safer and they streamline your organisation.

Take part and have influence

As a member of SIS you will have the possibility to participate in standardization activities on national, European and global level. The membership in SIS will give you the opportunity to influence future standards and gain access to early stage information about developments within your field.

Get to know the finished work

We offer our customers everything in connection with standards and their application. You can purchase all the publications you need from us - everything from individual standards, technical reports and standard packages through to manuals and online services. Our web service e-nav gives you access to an easy-to-navigate library where all standards that are relevant to your company are available. Standards and manuals are sources of knowledge. We sell them.

Increase understanding and improve perception

With SIS you can undergo either shared or in-house training in the content and application of standards. Thanks to our proximity to international development and ISO you receive the right knowledge at the right time, direct from the source. With our knowledge about the potential of standards, we assist our customers in creating tangible benefit and profitability in their organisations.

If you want to know more about SIS, or how standards can streamline your organisation, please visit www.sis.se or contact us on phone +46 (0)8-555 523 00



Europastandarden EN ISO 14253-5:2015 gäller som svensk standard. Detta dokument innehåller den officiella engelska versionen av EN ISO 14253-5:2015.

The European Standard EN ISO 14253-5:2015 has the status of a Swedish Standard. This document contains the official English version of EN ISO 14253-5:2015.

© Copyright/Upphovsrätten till denna produkt tillhör SIS, Swedish Standards Institute, Stockholm, Sverige. Användningen av denna produkt regleras av slutanvändarlicensen som återfinns i denna produkt, se standardens sista sidor.

© Copyright SIS, Swedish Standards Institute, Stockholm, Sweden. All rights reserved. The use of this product is governed by the end-user licence for this product. You will find the licence in the end of this document.

Upplysningar om sakinnehållet i standarden lämnas av SIS, Swedish Standards Institute, telefon 08-555 520 00. Standarder kan beställas hos SIS Förlag AB som även lämnar allmänna upplysningar om svensk och utländsk standard.

Information about the content of the standard is available from the Swedish Standards Institute (SIS), telephone +46 8 555 520 00. Standards may be ordered from SIS Förlag AB, who can also provide general information about Swedish and foreign standards.

Denna standard är framtagen av kommittén för Mätteknik GPS och Ytstruktur, SIS/TK 507/AG 6.

Har du synpunkter på innehållet i den här standarden, vill du delta i ett kommande revideringsarbete eller vara med och ta fram andra standarder inom området? Gå in på www.sis.se - där hittar du mer information.

Contents

	Page
Foreword	iv
Introduction	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	2
4 General	6
5 Test measurand	7
5.1 General.....	7
5.2 Input quantities and test measurand definition.....	8
6 Tester responsibility criterion	9
7 Specific issues in testing indicating measuring instruments	11
7.1 General.....	11
7.2 Errors of the indicating measuring instrument.....	11
7.3 Errors in user-provided quantity values.....	12
7.4 Using alternative test equipment.....	12
Annex A (informative) Guidance on using alternative test equipment	14
Annex B (informative) Relation to the GPS matrix model	16
Bibliography	17

Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

The procedures used to develop this document and those intended for its further maintenance are described in the ISO/IEC Directives, Part 1. In particular the different approval criteria needed for the different types of ISO documents should be noted. This document was drafted in accordance with the editorial rules of the ISO/IEC Directives, Part 2 (see www.iso.org/directives).

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights. Details of any patent rights identified during the development of the document will be in the Introduction and/or on the ISO list of patent declarations received (see www.iso.org/patents).

Any trade name used in this document is information given for the convenience of users and does not constitute an endorsement.

For an explanation on the meaning of ISO specific terms and expressions related to conformity assessment, as well as information about ISO's adherence to the WTO principles in the Technical Barriers to Trade (TBT) see the following URL: [Foreword - Supplementary information](#)

The committee responsible for this document is ISO/TC 213, *Dimensional and geometrical product specifications and verifications*.

ISO 14253 consists of the following parts, under the general title *Geometrical product specifications (GPS) — Inspection by measurement of workpieces and measuring equipment*:

- *Part 1: Decision rules for proving conformity or nonconformity with specifications*
- *Part 2: Guide to the estimation of uncertainty in GPS measurement, in calibration of measuring equipment and in product verification*
- *Part 3: Guidelines for achieving agreements on measurement uncertainty statements*
- *Part 4: Background on functional limits and specification limits in decision rules*
- *Part 5: Uncertainty in verification testing of indicating measuring instruments*
- *Part 6: Generalized decision rules for the acceptance and rejection of instruments and workpieces*
[Technical Report]

Introduction

This part of ISO 14253 belongs to the general geometrical product specification (GPS) series of documents (see ISO 14638). It influences chain link F of all chains of standards in the general GPS matrix.

The ISO/GPS matrix model given in ISO 14638 gives an overview of the ISO/GPS system of which this international standard is a part. The fundamental rules of ISO/GPS given in ISO 8015 apply to this part of ISO 14253 and the default decision rules given in ISO 14253-1 apply to specifications made in accordance with this part of ISO 14253, unless otherwise indicated.

For more detailed information about the relationship of this part of ISO 14253 to other standards and to the GPS matrix model, see [Annex B](#).

Decision rules for deciding conformity or non-conformity to specifications are based on the measurement uncertainty incurred while testing.

Usual practice in measurement familiarizes metrologists and practitioners with measurement uncertainty. Any possible effect that may affect the measurement result is considered and quantified as an uncertainty component and is eventually included in the combined uncertainty. The purpose of the measurement is to gather quantitative information on a given measurand, and the uncertainty statement expresses how reliable that information is.

In the case of tests of indicating measuring instruments, the purpose of the measurement is to investigate one or more metrological characteristics of the indicating measuring instrument rather than to measure characteristics of features of a workpiece. The uncertainty being evaluated in this case, the test value uncertainty, quantifies the accuracy of the test value. The test detects the quality of the indicating measuring instrument, reported through the test values and not through the test value uncertainty.

The test value uncertainty for indicating measuring instruments is not conceptually trivial to evaluate, and careful consideration is necessary to determine which uncertainty components should and which should not be accounted for.

Some tests of indicating measuring instruments may be relative to quantities other than instrument indications, or a single test may investigate both the instrument indication(s) and other metrological characteristics. An example is a test of a micrometer investigating the indication error (subject to an MPE) as well as the measuring force (subject to an MPL). For tests, or portions of them, relative to metrological characteristics other than instrument indications, this part of ISO 14253 is not applicable: they are about quantities for which the application of the ISO/IEC Guide 98-3 (GUM) and of the ISO 14253-2 is conceptually straightforward, with no need of further guidance in this part of ISO 14253.

A rigorous definition of the test value uncertainty when testing indicating measuring instruments is given. Application of conventional uncertainty evaluation based on this definition and according to the ISO/IEC Guide 98-3 (GUM) and the ISO 14253-2 determines which uncertainty components to account for.

Geometrical product specifications (GPS) — Inspection by measurement of workpieces and measuring equipment —

Part 5: Uncertainty in verification testing of indicating measuring instruments

1 Scope

This part of ISO 14253 specifies concepts and terms for evaluating the uncertainties of the test values derived according to a test protocol agreed upon by the parties and relative to instrument indication(s), obtained in verification testing of GPS indicating measuring instruments.

NOTE The uncertainty of the test values, referred to as test value uncertainty, is not to be confused with the measurement uncertainty associated with using that indicating measuring instrument to measure workpieces. The former only is covered in this part of ISO 14253; for guidance on the latter see the ISO/IEC Guide 98-3 (GUM) and ISO 14253-2.

When a test of an indicating measuring instrument comprises several test values, some relative to the instrument indication and some to other metrological characteristics, this part of ISO 14253 is concerned with the uncertainty of the former only.

This part of ISO 14253 does not provide guidelines to ensure the adequacy of a test protocol; rather, once a test protocol is given, it describes how to evaluate the consequent test value uncertainty.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 10360-1:2000, *Geometrical Product Specifications (GPS) — Acceptance and reverification tests for coordinate measuring machines (CMM) — Part 1: Vocabulary*

ISO 14253-1:2013, *Geometrical product specifications (GPS) — Inspection by measurement of workpieces and measuring equipment — Part 1: Decision rules for proving conformity or nonconformity with specifications*

ISO/TR 14253-6:2012, *Geometrical product specifications (GPS) — Inspection by measurement of workpieces and measuring equipment — Part 6: Generalized decision rules for the acceptance and rejection of instruments and workpieces*

ISO 14978:2006, *Geometrical product specifications (GPS) — General concepts and requirements for GPS measuring equipment*

ISO 17450-2:2012, *Geometrical product specifications (GPS) — General concepts — Part 2: Basic tenets, specifications, operators, uncertainties and ambiguities*

ISO/IEC Guide 98-3:2008, *Uncertainty of measurement — Part 3: Guide to the expression of uncertainty in measurement (GUM:1995)*

ISO/IEC Guide 99:2007, *International vocabulary of metrology — Basic and general concepts and associated terms (VIM)*

3 Terms and definitions

For the purpose of this document, the terms and definitions given in ISO 10360-1, ISO 14253-1, ISO/TR 14253-6, ISO 14978, ISO 17450-2, ISO/IEC Guide 98-3 (GUM), ISO/IEC Guide 99 (VIM), and the following apply.

3.1

test

<of a GPS indicating measuring instrument> sequence of preparatory, measurement, mathematical and decisional actions according to a test protocol

Note 1 to entry: Not all steps in the sequence are necessarily present in a test protocol.

Note 2 to entry: Tests are often used to verify the specifications of a GPS indicating measuring instrument.

Note 3 to entry: The specification of an indicating measuring instrument may be expressed by one or more MPEs (Maximum Permissible Errors).

Note 4 to entry: Prominent cases of tests are the acceptance test and the reverification test.

Note 5 to entry: This term is sometimes used in a wider sense, encompassing cases when a test produces a binary or categorical result. An example of a binary assessment is determining whether or not a software algorithm converged. For the purpose of this part of ISO 14253, tests are restricted to those based on test values.

Note 6 to entry: See [Figure 1](#).

3.2

test instance

combination of test equipment, set up, measurement sequence, environmental and instrumental conditions of a test, which yields a test value(s)

3.3

permissible test instance

test instance in compliance with the test protocol, and with the alternatives and stipulations therein

Note 1 to entry: An alternative occurs when the test protocol allows options, either discrete among enumerated cases, or continuous in a range of permissible values. An example of the former is the choice of the test equipment, e.g. a gauge block or a step gauge for testing a CMM; an example of the latter is the ambient temperature within the required test conditions.

Note 2 to entry: A stipulation occurs when the test protocol specifies the amount of measurement in a test, e.g. a specific number of repeated measurements.

Note 3 to entry: A test may be subject to alternatives and stipulations at the same time. For instance, test equipment is applied to an indicating measuring instrument in a limited number of configurations (stipulation) chosen at the tester counterpart's discretion (alternative).

Note 4 to entry: Alternatives serve two purposes. (1) To accommodate to actual conditions. For instance, alternative test equipment to accommodate actual availability, or any environmental condition within the required test conditions to accommodate the actual testing environment. (2) To leave details of the test unspecified up to the time of testing, to encourage the indicating measuring instrument manufacturer – in order to avoid non-acceptance of the instrument – to deliver overall compliant indicating measuring instruments. For instance, some procedural details may be left to the *tester counterpart* ([3.14](#)) to decide at the time of testing, to force the manufacturer to deliver a compliant indicating measuring instrument for any possible procedural option.

3.4

test measurand

metrological characteristic of an indicating measuring instrument intended to be verified in a test, based on a single permissible test instance, defined by a test protocol

Note 1 to entry: A test protocol may allow for multiple permissible test instances, to adapt to actual circumstances and to limit the experimental effort. The test measurand is defined for each single test instance, and different permissible test instances may give raise to different test measurands.

3.5 test protocol

predefined detailed specification of a test which defines the test measurand, the required test conditions and a decision rule

Note 1 to entry: The test protocol is defined either by relevant standards or – when none is available – by the tester or the *tester counterpart* (3.14).

Note 2 to entry: The tester and the tester counterpart are to agree upon the test protocol prior to the test.

Note 3 to entry: A default decision rule is given in ISO 14253-1. See ISO/TR 14253-6 for guidance in defining alternative decision rules.

Note 4 to entry: An unambiguous test protocol is crucial for the effectiveness of a test. In particular, the definition of the set of permissible test instances constitutes a trade-off between thoroughness and practical and economical viability of the test.

Note 5 to entry: As the default rule in ISO 14253-1 is stringent and conservative, in this case the verification approaches a proof in an absolute sense.

3.6 measured test indication

result of a measurement performed in a test, which contributes to the test value according to a test operator

Note 1 to entry: A test value may be based either on a single or on multiple measured test indications, as stipulated in the test protocol.

3.7 test operator

predefined sequence of mathematical and/or statistical operations applied to the measured test indication(s) collected in the test to deliver a test value

Note 1 to entry: Each test value is delivered according to a test operator. In the case of a test yielding multiple test values (see 3.8 Note 4 to entry), as many test operators are needed.

Note 2 to entry: The operations in the sequence can be divided in four broad categories: outlier rejection, noise reduction, statistics, and other mathematical operations.

- Examples of outlier rejection: (1) discarding measured test indications above the 99th percentile of the measured test indications collected in the test; (2) when no more than 2 % of the measured test indications causes failure to meet the specification, repeating those measurement three times.
- Examples of noise reduction: (1) taking the median of repeated measurement values; (2) performing a (spatial) frequency analysis and discarding all wavelengths above a predefined threshold.
- Examples of statistics: taking (1) the mean or (2) the maximum of the absolute values of the measured test indications collected in the test.
- Examples of other mathematical operations: (1) computing a Gaussian (best-fit) sphere associated to the coordinates obtained as measured test indications, and the individual distances of each measured test indication to the sphere centre; (2) computing the mean of the measured test indications taken in a scan over a line.

3.8 test value

quantity value measured in a test estimating the magnitude of a test measurand

Note 1 to entry: A test value is based on the measured test indication(s) and is derived according to the test operator.

Note 2 to entry: A test value cannot usually capture the performance of an indicating measuring instrument in full, because is limited as regulated by the test protocol.

Note 3 to entry: A test value may be derived from several measured test indications, according to the test operator.