

# SVENSK STANDARD

## SS-EN ISO 7029:2017



Fastställt/Approved: 2017-02-08  
Publicerad/Published: 2017-02-13  
Utgåva/Edition: 2  
Språk/Language: engelska/English  
ICS: 13.140

---

**Akustik – Statistisk fördelning av hörtrösklar som en funktion av ålder och kön (ISO 7029:2017)**

**Acoustics – Statistical distribution of hearing thresholds related to age and gender (ISO 7029:2017)**

This preview is downloaded from [www.sis.se](http://www.sis.se). Buy the entire standard via <https://www.sis.se/std-8024963>

# Standarder får världen att fungera

*SIS (Swedish Standards Institute) är en fristående ideell förening med medlemmar från både privat och offentlig sektor. Vi är en del av det europeiska och globala nätverk som utarbetar internationella standarder. Standarder är dokumenterad kunskap utvecklad av framstående aktörer inom industri, näringsliv och samhälle och befrämjar handel över gränser, bidrar till att processer och produkter blir säkrare samt effektiviserar din verksamhet.*

## Delta och påverka

Som medlem i SIS har du möjlighet att påverka framtida standarder inom ditt område på nationell, europeisk och global nivå. Du får samtidigt tillgång till tidig information om utvecklingen inom din bransch.

## Ta del av det färdiga arbetet

Vi erbjuder våra kunder allt som rör standarder och deras tillämpning. Hos oss kan du köpa alla publikationer du behöver – allt från enskilda standarder, tekniska rapporter och standardpaket till handböcker och onlinetjänster. Genom vår webbtjänst e-nav får du tillgång till ett lättnavigerat bibliotek där alla standarder som är aktuella för ditt företag finns tillgängliga. Standarder och handböcker är källor till kunskap. Vi säljer dem.

## Utveckla din kompetens och lyckas bättre i ditt arbete

Hos SIS kan du gå öppna eller företagsinterna utbildningar kring innehåll och tillämpning av standarder. Genom vår närhet till den internationella utvecklingen och ISO får du rätt kunskap i rätt tid, direkt från källan. Med vår kunskap om standarders möjligheter hjälper vi våra kunder att skapa verklig nytta och lönsamhet i sina verksamheter.

**Vill du veta mer om SIS eller hur standarder kan effektivisera din verksamhet är du välkommen in på [www.sis.se](http://www.sis.se) eller ta kontakt med oss på tel 08-555 523 00.**



# Standards make the world go round

*SIS (Swedish Standards Institute) is an independent non-profit organisation with members from both the private and public sectors. We are part of the European and global network that draws up international standards. Standards consist of documented knowledge developed by prominent actors within the industry, business world and society. They promote cross-border trade, they help to make processes and products safer and they streamline your organisation.*

## Take part and have influence

As a member of SIS you will have the possibility to participate in standardization activities on national, European and global level. The membership in SIS will give you the opportunity to influence future standards and gain access to early stage information about developments within your field.

## Get to know the finished work

We offer our customers everything in connection with standards and their application. You can purchase all the publications you need from us - everything from individual standards, technical reports and standard packages through to manuals and online services. Our web service e-nav gives you access to an easy-to-navigate library where all standards that are relevant to your company are available. Standards and manuals are sources of knowledge. We sell them.

## Increase understanding and improve perception

With SIS you can undergo either shared or in-house training in the content and application of standards. Thanks to our proximity to international development and ISO you receive the right knowledge at the right time, direct from the source. With our knowledge about the potential of standards, we assist our customers in creating tangible benefit and profitability in their organisations.

**If you want to know more about SIS, or how standards can streamline your organisation, please visit [www.sis.se](http://www.sis.se) or contact us on phone +46 (0)8-555 523 00**



Europastandarden EN ISO 7029:2017 gäller som svensk standard. Detta dokument innehåller den officiella engelska versionen av EN ISO 7029:2017.

Denna standard ersätter SS-EN ISO 7029, utgåva 1.

The European Standard EN ISO 7029:2017 has the status of a Swedish Standard. This document contains the official English version of EN ISO 7029:2017.

This standard supersedes the Swedish Standard SS-EN ISO 7029, edition 1.

© Copyright/Upphovsrätten till denna produkt tillhör SIS, Swedish Standards Institute, Stockholm, Sverige. Användningen av denna produkt regleras av slutanvändarlicensen som återfinns i denna produkt, se standardens sista sidor.

© Copyright SIS, Swedish Standards Institute, Stockholm, Sweden. All rights reserved. The use of this product is governed by the end-user licence for this product. You will find the licence in the end of this document.

*Uppllysningar om sakinnehållet i standarden lämnas av SIS, Swedish Standards Institute, telefon 08-555 520 00. Standarder kan beställas hos SIS Förlag AB som även lämnar allmänna uppllysningar om svensk och utländsk standard.*

*Information about the content of the standard is available from the Swedish Standards Institute (SIS), telephone +46 8 555 520 00. Standards may be ordered from SIS Förlag AB, who can also provide general information about Swedish and foreign standards.*

Denna standard är framtagen av kommittén för Akustik och buller, SIS/TK 110.

Har du synpunkter på innehållet i den här standarden, vill du delta i ett kommande revideringsarbete eller vara med och ta fram andra standarder inom området? Gå in på [www.sis.se](http://www.sis.se) - där hittar du mer information.



EUROPEAN STANDARD

**EN ISO 7029**

NORME EUROPÉENNE

EUROPÄISCHE NORM

January 2017

ICS 13.140

Supersedes EN ISO 7029:2000

English Version

## Acoustics - Statistical distribution of hearing thresholds related to age and gender (ISO 7029:2017)

Acoustique - Distribution statistique des seuils  
d'audition en fonction de l'âge et du sexe (ISO  
7029:2017)

This European Standard was approved by CEN on 14 January 2017.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION  
COMITÉ EUROPÉEN DE NORMALISATION  
EUROPÄISCHES KOMITEE FÜR NORMUNG

**CEN-CENELEC Management Centre: Avenue Marnix 17, B-1000 Brussels**

# Contents

Page

<b>European foreword</b> .....	<b>☞</b>
<b>Introduction</b> .....	<b>v</b>
<b>1 Scope</b> .....	<b>1</b>
<b>2 Normative references</b> .....	<b>2</b>
<b>3 Terms and definitions</b> .....	<b>2</b>
<b>4 Specification</b> .....	<b>2</b>
4.1 General.....	2
4.2 Median.....	3
4.3 Distribution around the median.....	3
4.4 Application of data.....	5
<b>Annex A (informative) Selected values of the Gaussian distribution</b> .....	<b>7</b>
<b>Annex B (informative) Numerical example to illustrate the procedure</b> .....	<b>8</b>
<b>Annex C (informative) Median values of expected hearing threshold deviations</b> .....	<b>9</b>
<b>Annex D (informative) Selected values of the statistical distribution of hearing threshold deviations</b> .....	<b>10</b>
<b>Annex E (informative) Expected median thresholds at audiometric frequencies from 9 000 Hz to 12 500 Hz</b> .....	<b>13</b>
<b>Annex F (informative) Notes on the derivation of descriptive statistics of hearing thresholds</b> .....	<b>15</b>
<b>Annex G (informative) Dispersion of source data around the expected median of hearing thresholds</b> .....	<b>20</b>
<b>Bibliography</b> .....	<b>21</b>

## **European foreword**

This document (EN ISO 7029:2017) has been prepared by Technical Committee ISO/TC 43 “Acoustics” in collaboration with Technical Committee CEN/TC 211 “Acoustics” the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by July 2017, and conflicting national standards shall be withdrawn at the latest by July 2017.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 7029:2000.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Serbia, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

### **Endorsement notice**

The text of ISO 7029:2017 has been approved by CEN as EN ISO 7029:2017 without any modification.

## **Introduction**

The sensitivity of human hearing is well known to decrease with age and the impairment of hearing develops more rapidly for sound at high frequencies than at low frequencies. Moreover, the magnitude of this effect varies considerably among individuals.

When testing the hearing of persons markedly over 18 years of age, part of any observed hearing loss will probably be associated with age. It is important to be aware of this when estimating the amount of hearing loss attributable to other causes under investigation.

It should be noted that a decrease in hearing ability may not necessarily be caused by ageing itself, but by many injurious influences during lifetime, which are not known in detail.

This document is based on a thorough examination of literature data on the differences between groups having different ages for populations of otologically normal persons as defined herein. Distinction is made between males and females since the difference is found to be of significance in the case of older age groups. The data have been derived from investigations using pure tones transmitted to the ear from an earphone, but no evidence is known that disqualifies their use for noise band stimuli.

This document is a revision of the second edition (ISO 7029:2000). The expected medians and statistical distributions of hearing thresholds were re-estimated using audiometric data published after the establishment of the first edition (ISO 7029:1984). All the data on which the second edition had been based were discarded. Thus, this third edition describes the hearing sensitivity profile of people in recent years.

Hearing thresholds presented in this document are generally lower at high frequencies than those in the previous editions of this document. The 4 kHz dip observed in males has become negligibly small. The source data of the previous editions might not have been screened rigorously in terms of hearing abnormalities. Problems related to instrumentation might also have affected measurement data.

The expected median hearing thresholds at the frequencies from 9 000 Hz to 12 500 Hz are presented for information. Audiometry at those frequencies is executable using an extended high-frequency audiometer.



# Acoustics — Statistical distribution of hearing thresholds related to age and gender

## 1 Scope

This document provides descriptive statistics of the hearing threshold deviation for populations of otologically normal persons of various ages under monaural earphone listening conditions. It specifies the following, for populations within the age limits from 18 years to 80 years for the range of audiometric frequencies from 125 Hz to 8 000 Hz:

- a) the expected median value of hearing thresholds given relative to the median hearing threshold at the age of 18 years;
- b) the expected statistical distribution above and below the median value.

For the frequencies from 3 000 Hz to 8 000 Hz, the median and statistical distribution for populations above 70 years are presented for information only.

This document also provides for information the expected median values at audiometric frequencies from 9 000 Hz to 12 500 Hz within the age limits from 22 years to 80 years.

The data are applicable for estimating the amount of hearing loss caused by a specific agent in a population. Such a comparison is valid if the population under study consists of persons who are otologically normal except for the effect of the specific agent. Noise exposure is an example of a specific agent and for this application, selected data from this document are referred to as “database A” in ISO 1999.

NOTE 1 ISO 1999:2013, Database A is based on a previous edition of ISO 7029.

The data may also be used to assess an individual’s hearing in relation to the distribution of hearing thresholds which is normal for the person’s age group. However, it is not possible to determine for an individual precisely which part of an observed hearing loss is attributable to an accumulation of detrimental effects on the hearing which increase with age, and which part has been caused by other factors such as noise.

The hearing threshold deviation as defined herein and the hearing threshold level as defined in other International Standards (ISO 389-1, ISO 389-2, ISO 389-5, ISO 389-8, ISO 8253-1, ISO 8253-2, IEC 60645-1) express the hearing threshold of an individual or an individual ear, respectively, relative to

- the expected median hearing threshold of 18-year-old age group of the same gender, or
- a reference zero level specified in various parts of ISO 389.

To the extent that the reference zero level represents the median of the 18-year-old population, the values of the two terms will be the same.

NOTE 2 The values of these two are not always the same for some reasons. One reason is that the reference zero level has been determined based on the hearing threshold levels of persons older than 18 years, including those aged up to 25 years or to 30 years, who have slightly worse hearing sensitivity on average.

NOTE 3 ISO 28961 specifies the expected statistical distribution of hearing thresholds, expressed in sound pressure level in decibels, for populations of otologically normal persons of the age from 18 years to 25 years under binaural, free-field listening conditions. It enables the calculation not only at audiometric frequencies, but also for other frequencies at one-third-octave intervals from 20 Hz to 16 000 Hz.

## SS-EN ISO 7029:2017 (E)

### 2 Normative references

There are no normative references in this document.

### 3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

ISO and IEC maintain terminological databases for use in standardization at the following addresses:

- IEC Electropedia: available at <http://www.electropedia.org/>
- ISO Online browsing platform: available at <https://www.iso.org/obp/>

**3.1**  
**otologically normal person**  
person in a normal state of health who is free from all signs or symptoms of ear disease and from obstructing wax in the ear canal and who has no history of undue exposure to noise, exposure to potentially ototoxic substances or familial hearing loss

[SOURCE: ISO 8253-1:2010, 3.7, modified — The term “ototoxic drugs” has been changed to “ototoxic substances”.]

**3.2**  
**hearing threshold deviation**  
 $\Delta H$   
threshold of hearing of an individual minus the median threshold of hearing of a population of 18-year-old *otologically normal persons* (3.1) of the same gender

Note 1 to entry: The term threshold of hearing is defined in ISO 389-1:1998, 3.4.

### 4 Specification

#### 4.1 General

The statistical distribution of hearing threshold deviations for otologically normal persons of a specific age  $Y$  and a specific gender is given in terms of the median value (see 4.2) and the distribution around the median (see 4.3).

The range of  $Y$  is from 18 years to 80 years for audiometric frequencies from 125 Hz to 8 000 Hz. For the frequencies from 3 000 Hz to 8 000 Hz, the  $Y$  value above 70 years is given for information purposes only.

NOTE The statistical distribution estimation for the age above 70 years at 3 000 Hz and above is subject to a large uncertainty because the hearing threshold level from many subjects is not obtainable (“scales out”) at those frequencies.

[Annex E](#) presents expected median thresholds at audiometric frequencies from 9 000 Hz to 12 500 Hz.

[Annex F](#) discusses the derivation of the descriptive statistics of hearing thresholds.

[Annex G](#) shows root-mean-square error (RMSE) values and estimating the uncertainty of the statistical values.

## 4.2 Median

The median value  $\Delta H_{\text{md},Y}$ , in decibels, at audiometric frequencies from 125 Hz to 8 000 Hz is given by [Formula \(1\)](#):

$$\Delta H_{\text{md},Y} = \alpha_{\text{md}} (Y - 18)^{\beta_{\text{md}}} \quad (1)$$

where

$\alpha_{\text{md}}$  and  $\beta_{\text{md}}$  are dimensionless quantities as given in [Table 1](#);

$Y$  is the age in years.

The age of 18 years is the lower limit of the  $Y$  range for which [Formula \(1\)](#) is valid.

**Table 1 — Values of  $\alpha_{\text{md}}$  and  $\beta_{\text{md}}$  in [Formula \(1\)](#)**

Frequency Hz	$\alpha_{\text{md}}$		$\beta_{\text{md}}$	
	Males	Females	Males	Females
125	$2,50 \times 10^{-6}$	$6,16 \times 10^{-4}$	3,841	2,451
250	$1,39 \times 10^{-4}$	$3,98 \times 10^{-4}$	2,832	2,568
500	$4,59 \times 10^{-4}$	$2,61 \times 10^{-4}$	2,537	2,708
750	$5,70 \times 10^{-4}$	$2,25 \times 10^{-4}$	2,512	2,775
1 000	$7,02 \times 10^{-4}$	$2,21 \times 10^{-4}$	2,494	2,805
1 500	$1,09 \times 10^{-3}$	$2,53 \times 10^{-4}$	2,446	2,813
2 000	$1,56 \times 10^{-3}$	$3,12 \times 10^{-4}$	2,404	2,792
3 000	$2,54 \times 10^{-3}$	$4,88 \times 10^{-4}$	2,350	2,728
4 000	$3,40 \times 10^{-3}$	$7,37 \times 10^{-4}$	2,325	2,660
6 000	$4,53 \times 10^{-3}$	$1,47 \times 10^{-3}$	2,315	2,539
8 000	$5,06 \times 10^{-3}$	$2,53 \times 10^{-3}$	2,328	2,439

## 4.3 Distribution around the median

The distribution around the median shall be approximated by the upper and lower halves, respectively, of two Gaussian distributions, each with their standard deviation  $s_{\text{u}}$  and  $s_{\text{l}}$ , in decibels, given by [Formulae \(2\)](#) and [\(3\)](#):

$$s_{\text{u}} = \gamma_{0,\text{su}} + \gamma_{1,\text{su}} (Y - 18) + \gamma_{2,\text{su}} (Y - 18)^2 + \gamma_{3,\text{su}} (Y - 18)^3 + \gamma_{4,\text{su}} (Y - 18)^4 + \gamma_{5,\text{su}} (Y - 18)^5 \quad (2)$$

$$s_{\text{l}} = \gamma_{0,\text{sl}} + \gamma_{1,\text{sl}} (Y - 18) + \gamma_{2,\text{sl}} (Y - 18)^2 + \gamma_{3,\text{sl}} (Y - 18)^3 + \gamma_{4,\text{sl}} (Y - 18)^4 + \gamma_{5,\text{sl}} (Y - 18)^5 \quad (3)$$

The values of coefficients  $\gamma_{n,\text{su}}$  and  $\gamma_{n,\text{sl}}$  ( $n = 0, 1, 2, \dots, 5$ ) are given in [Tables 2](#) and [3](#) for males and in [Tables 4](#) and [5](#) for females. The age of 18 years is the lower limit of the  $Y$  range for which [Formulae \(2\)](#) and [\(3\)](#) are valid.