

# SVENSK STANDARD

## SS-ISO 20857:2011




Fastställt/Approved: 2011-12-15  
Publicerad/Published: 2011-12-19  
Utgåva/Edition: 1  
Språk/Language: engelska/English  
ICS: 11.080.01

---

**Sterilisering av medicintekniska produkter – Torr värme – Krav för utveckling, validering och rutinkontroll av steriliseringsprocesser för medicintekniska produkter (ISO 20857:2010, IDT)**

**Sterilization of health care products – Dry heat – Requirements for the development, validation and routine control of a sterilization process for medical devices (ISO 20857:2010, IDT)**



# Standarder får världen att fungera

*SIS (Swedish Standards Institute) är en fristående ideell förening med medlemmar från både privat och offentlig sektor. Vi är en del av det europeiska och globala nätverk som utarbetar internationella standarder. Standarder är dokumenterad kunskap utvecklad av framstående aktörer inom industri, näringsliv och samhälle och befrämjar handel över gränser, bidrar till att processer och produkter blir säkrare samt effektiviserar din verksamhet.*

## Delta och påverka

Som medlem i SIS har du möjlighet att påverka framtida standarder inom ditt område på nationell, europeisk och global nivå. Du får samtidigt tillgång till tidig information om utvecklingen inom din bransch.

## Ta del av det färdiga arbetet

Vi erbjuder våra kunder allt som rör standarder och deras tillämpning. Hos oss kan du köpa alla publikationer du behöver – allt från enskilda standarder, tekniska rapporter och standardpaket till handböcker och onlinetjänster. Genom vår webbtjänst e-nav får du tillgång till ett lättnavigerat bibliotek där alla standarder som är aktuella för ditt företag finns tillgängliga. Standarder och handböcker är källor till kunskap. Vi säljer dem.

## Utveckla din kompetens och lyckas bättre i ditt arbete

Hos SIS kan du gå öppna eller företagsinterna utbildningar kring innehåll och tillämpning av standarder. Genom vår närhet till den internationella utvecklingen och ISO får du rätt kunskap i rätt tid, direkt från källan. Med vår kunskap om standarders möjligheter hjälper vi våra kunder att skapa verklig nytta och lönsamhet i sina verksamheter.

**Vill du veta mer om SIS eller hur standarder kan effektivisera din verksamhet är du välkommen in på [www.sis.se](http://www.sis.se) eller ta kontakt med oss på tel 08-555 523 00.**



# Standards make the world go round

*SIS (Swedish Standards Institute) is an independent non-profit organisation with members from both the private and public sectors. We are part of the European and global network that draws up international standards. Standards consist of documented knowledge developed by prominent actors within the industry, business world and society. They promote cross-border trade, they help to make processes and products safer and they streamline your organisation.*

## Take part and have influence

As a member of SIS you will have the possibility to participate in standardization activities on national, European and global level. The membership in SIS will give you the opportunity to influence future standards and gain access to early stage information about developments within your field.

## Get to know the finished work

We offer our customers everything in connection with standards and their application. You can purchase all the publications you need from us - everything from individual standards, technical reports and standard packages through to manuals and online services. Our web service e-nav gives you access to an easy-to-navigate library where all standards that are relevant to your company are available. Standards and manuals are sources of knowledge. We sell them.

## Increase understanding and improve perception

With SIS you can undergo either shared or in-house training in the content and application of standards. Thanks to our proximity to international development and ISO you receive the right knowledge at the right time, direct from the source. With our knowledge about the potential of standards, we assist our customers in creating tangible benefit and profitability in their organisations.

**If you want to know more about SIS, or how standards can streamline your organisation, please visit [www.sis.se](http://www.sis.se) or contact us on phone +46 (0)8-555 523 00**



Den internationella standarden ISO 20857:2010 gäller som svensk standard. Detta dokument innehåller den officiella engelska versionen av ISO 20857:2010.

The International Standard ISO 20857:2010 has the status of a Swedish Standard. This document contains the official version of ISO 20857:2010.

© Copyright/Upphovsrätten till denna produkt tillhör SIS, Swedish Standards Institute, Stockholm, Sverige. Användningen av denna produkt regleras av slutanvändarlicensen som återfinns i denna produkt, se standardens sista sidor.

© Copyright SIS, Swedish Standards Institute, Stockholm, Sweden. All rights reserved. The use of this product is governed by the end-user licence for this product. You will find the licence in the end of this document.

*Upplysningar om sakinnehållet i standarden lämnas av SIS, Swedish Standards Institute, telefon 08-555 520 00. Standarder kan beställas hos SIS Förlag AB som även lämnar allmänna upplysningar om svensk och utländsk standard.*

*Information about the content of the standard is available from the Swedish Standards Institute (SIS), telephone +46 8 555 520 00. Standards may be ordered from SIS Förlag AB, who can also provide general information about Swedish and foreign standards.*

Denna standard är framtagen av kommittén för Sterilisering av medicintekniska produkter, SIS/TK 349.

Har du synpunkter på innehållet i den här standarden, vill du delta i ett kommande revideringsarbete eller vara med och ta fram andra standarder inom området? Gå in på [www.sis.se](http://www.sis.se) - där hittar du mer information.



# Contents

Page

Foreword .....	v
Introduction.....	vi
1 Scope .....	1
1.1 Inclusions .....	1
1.2 Exclusions.....	1
2 Normative references.....	2
3 Terms and definitions .....	2
4 Quality management system elements .....	10
4.1 Documentation .....	10
4.2 Management responsibility .....	10
4.3 Product realization .....	10
4.4 Measurement, analysis and improvement — Control of nonconforming product.....	10
5 Sterilizing agent characterization .....	11
5.1 Sterilizing agent.....	11
5.2 Microbicidal effectiveness.....	11
5.3 Material effects .....	11
5.4 Environmental considerations.....	11
6 Process and equipment characterization .....	11
6.1 Process characterization.....	11
6.2 Equipment characterization .....	11
7 Product definition.....	13
7.1 General .....	13
7.2 Product safety and performance .....	13
7.3 Packaging considerations .....	14
7.4 Microbiological quality.....	14
7.5 Product family.....	14
7.6 Biological safety .....	14
8 Process definition .....	15
9 Validation .....	16
9.1 General .....	16
9.2 Installation qualification .....	16
9.3 Operational qualification .....	16
9.4 Performance qualification .....	16
9.5 Additional sterilization systems .....	18
9.6 Review and approval of validation .....	18
10 Routine monitoring and control.....	19
10.1 Routine control .....	19
10.2 Routine monitoring .....	19
10.3 Process monitoring locations.....	20
11 Product release from sterilization/depyrogenation .....	21
12 Maintaining process effectiveness.....	21
12.1 General .....	21
12.2 Recalibration.....	21
12.3 Maintenance of equipment .....	21
12.4 Requalification.....	21
12.5 Assessment of change .....	22

<b>Annex A</b> (informative) <b>Guidance on the application of this International Standard</b> .....	<b>23</b>
<b>Annex B</b> (informative) <b>Process definition based on inactivation of the microbial population in its natural state (bioburden-based approach)</b> .....	<b>46</b>
<b>Annex C</b> (informative) <b>Process definition based on the inactivation of reference microorganisms and knowledge of bioburden (combined bioburden/biological indicator approach)</b> .....	<b>48</b>
<b>Annex D</b> (informative) <b>Conservative process definition based on inactivation of reference microorganisms (overkill method)</b> .....	<b>51</b>
<b>Annex E</b> (informative) <b>Process development</b> .....	<b>54</b>
<b>Bibliography</b> .....	<b>57</b>

## Foreword

ISO (the International Organization for Standardization) is a worldwide federation of national standards bodies (ISO member bodies). The work of preparing International Standards is normally carried out through ISO technical committees. Each member body interested in a subject for which a technical committee has been established has the right to be represented on that committee. International organizations, governmental and non-governmental, in liaison with ISO, also take part in the work. ISO collaborates closely with the International Electrotechnical Commission (IEC) on all matters of electrotechnical standardization.

International Standards are drafted in accordance with the rules given in the ISO/IEC Directives, Part 2.

The main task of technical committees is to prepare International Standards. Draft International Standards adopted by the technical committees are circulated to the member bodies for voting. Publication as an International Standard requires approval by at least 75 % of the member bodies casting a vote.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. ISO shall not be held responsible for identifying any or all such patent rights.

ISO 20857 was prepared by Technical Committee ISO/TC 198, *Sterilization of health care products*.

## Introduction

A sterile medical device is one that is free of viable microorganisms. International Standards that specify requirements for development, validation and routine control of sterilization processes, require, when it is necessary to supply a sterile medical device, that adventitious microbiological contamination of a medical device prior to sterilization be minimized. Even so, medical devices produced under standard manufacturing conditions in accordance with the requirements for quality management systems (see, for example, ISO 13485) may, prior to sterilization, have microorganisms on them, albeit in low numbers. Such products are non-sterile. The purpose of sterilization is to inactivate the microbiological contaminants and thereby transform the non-sterile products into sterile ones.

The kinetics of inactivation of a pure culture of microorganisms by physical and/or chemical agents used to sterilize medical devices can generally best be described by an exponential relationship between the numbers of microorganisms surviving and the extent of treatment with the sterilizing agent; inevitably this means that there is always a finite probability that a microorganism may survive regardless of the extent of treatment applied. For a given treatment, the probability of survival is determined by the number and resistance of microorganisms and by the environment in which the organisms exist during treatment. It follows that the sterility of any one product in a population subjected to sterilization processing cannot be guaranteed and the sterility of a processed population is defined in terms of the probability of there being a viable microorganism present on a product.

This International Standard describes requirements that, if met, will provide a dry heat sterilization process capable of sterilizing medical devices through appropriate microbicidal activity. This International Standard also describes requirements that, if met, will provide a dry heat depyrogenation process through an appropriate denaturation activity. Furthermore, such compliance permits prediction, with reasonable confidence, that there is a low probability of there being a viable microorganism present on the product after processing. Specification of this probability is a matter for regulatory authorities and may vary from country to country (see for example EN 556-1 and ANSI/AAMI ST67). Additionally, there will be a low probability of pyrogenic material of bacterial origin being present on the product after the application of a depyrogenation process.

Generic requirements of the quality management systems for design/development, production, installation and servicing are given in ISO 9001 and particular requirements for quality management systems for medical device production in ISO 13485. The standards for quality management systems recognise that, for certain processes used in manufacturing or reprocessing, the effectiveness of the process cannot be fully verified by subsequent inspection and testing of the product. Sterilization and depyrogenation are examples of such processes. For this reason, sterilization and depyrogenation processes are validated for use, the performance of the processes is monitored routinely, and the equipment is maintained.

Exposure to a properly validated, accurately controlled sterilization process is not the only factor associated with the provision of reliable assurance that the product is sterile and, in this regard, suitable for its intended use. Attention is therefore given to a number of factors including:

- a) the microbiological status of incoming raw materials and/or components;
- b) the validation and routine control of any cleaning and disinfection procedures used on the product;
- c) the control of the environment in which the product is manufactured, assembled and packaged;
- d) the control of equipment and processes;
- e) the control of personnel and their hygiene;
- f) the manner and materials in which the product is packaged;
- g) the conditions under which product is stored.



These factors also need consideration for the provision of reliable assurance of depyrogenation.

The type of contamination on the product to be sterilized varies and this variation influences the effectiveness of a sterilization and depyrogenation process. Product that has been used in a health care setting and is being presented for resterilization in accordance with the manufacturer's instructions (see ISO 17664) should be regarded as a special case. There is potential for such product to possess a wide range of contaminating microorganisms and residual inorganic and/or organic contamination in spite of the application of a cleaning process. Hence, particular attention has to be given to the validation and control of the cleaning and disinfection processes used during reprocessing.

The requirements are the normative parts of this International Standard with which compliance is claimed. The guidance given in the informative annexes is not normative and is not provided as a check list for auditors. The guidance provides explanations as well as methods that are accepted as being suitable means for complying with the requirements. Approaches other than those given in the guidance may be used if they are effective in achieving compliance with the requirements of this International Standard.

The development, validation and routine control of a sterilization process and/or a depyrogenation process comprise a number of discrete but interrelated activities, for example calibration, maintenance, product definition, process definition, installation qualification, operational qualification and performance qualification. While the activities required by this International Standard have been grouped together and are presented in a particular order, this International Standard does not require that the activities be performed in the order that they are presented. The activities required are not necessarily sequential, as the programmes of development and validation might be iterative. It is possible that performing these different activities will involve a number of separate individuals and/or organizations, each of whom undertake one or more of these activities. This International Standard does not specify the particular individuals or organizations to carry out the activities.



# Sterilization of health care products — Dry heat — Requirements for the development, validation and routine control of a sterilization process for medical devices

## 1 Scope

### 1.1 Inclusions

**1.1.1** This International Standard specifies requirements for the development, validation and routine control of a dry heat sterilization process for medical devices.

NOTE Although the scope of this International Standard is limited to medical devices, it specifies requirements and provides guidance that might be applicable to other health care products.

**1.1.2** Although this International Standard primarily addresses dry heat sterilization, it also specifies requirements and provides guidance in relation to depyrogenation processes using dry heat.

NOTE Dry heat is often used for the depyrogenation of equipment, components and health care products and its effectiveness has been demonstrated. The process parameters for sterilization and/or depyrogenation are time and temperature. Because the conditions for depyrogenation are typically more severe than those required for sterilization, a process that has been validated for product depyrogenation will result in product sterility without additional validation.

### 1.2 Exclusions

**1.2.1** This International Standard does not specify requirements for the development, validation and routine control of a process for inactivating the causative agents of spongiform encephalopathies such as scrapie, bovine spongiform encephalopathy and Creutzfeldt-Jakob disease.

NOTE See also ISO 22442-1, ISO 22442-2 and ISO 22442-3.

**1.2.2** This International Standard does not apply to processes that use infrared or microwaves as the heating technique.

**1.2.3** This International Standard does not detail a specified requirement for designating a medical device as "sterile."

NOTE Attention is drawn to national or regional requirements for designating medical devices as "sterile." See, for example, EN 556-1 or ANSI/AAMI ST67.

**1.2.4** This International Standard does not specify a quality management system for the control of all stages of production of medical devices.

NOTE It is not a requirement of this International Standard to have a complete quality management system during manufacture, but the elements of a quality management system that are the minimum necessary to control the sterilization process are normatively referenced at appropriate places in the text (see, in particular, Clause 4). Attention is drawn to the standards for quality management systems (see ISO 13485) that control all stages of production of medical devices, including the sterilization process. Regional and national regulations for the provision of medical devices might require implementation of a complete quality management system and the assessment of that system by a third party.

**1.2.5** This International Standard does not specify requirements for occupational safety associated with the design and operation of dry heat sterilization and/or depyrogenation facilities.