

SVENSK STANDARD

SS-EN 16280:2012



Fastställt/Approved: 2012-10-16
Publicerad/Published: 2012-10-17
Utgåva/Edition: 1
Språk/Language: engelska/English
ICS: 13.320; 71.040.40

Alkoholmätare för utandningskontroll avsedd för allmänt bruk – Krav och provningsmetoder

Breath alcohol test devices for general public – Requirements and test methods

This preview is downloaded from www.sis.se. Buy the entire standard via <https://www.sis.se/std-87753>

Standarder får världen att fungera

SIS (Swedish Standards Institute) är en fristående ideell förening med medlemmar från både privat och offentlig sektor. Vi är en del av det europeiska och globala nätverk som utarbetar internationella standarder. Standarder är dokumenterad kunskap utvecklad av framstående aktörer inom industri, näringsliv och samhälle och befrämjar handel över gränser, bidrar till att processer och produkter blir säkrare samt effektiviserar din verksamhet.

Delta och påverka

Som medlem i SIS har du möjlighet att påverka framtida standarder inom ditt område på nationell, europeisk och global nivå. Du får samtidigt tillgång till tidig information om utvecklingen inom din bransch.

Ta del av det färdiga arbetet

Vi erbjuder våra kunder allt som rör standarder och deras tillämpning. Hos oss kan du köpa alla publikationer du behöver – allt från enskilda standarder, tekniska rapporter och standardpaket till handböcker och onlinetjänster. Genom vår webbtjänst e-nav får du tillgång till ett lättnavigerat bibliotek där alla standarder som är aktuella för ditt företag finns tillgängliga. Standarder och handböcker är källor till kunskap. Vi säljer dem.

Utveckla din kompetens och lyckas bättre i ditt arbete

Hos SIS kan du gå öppna eller företagsinterna utbildningar kring innehåll och tillämpning av standarder. Genom vår närhet till den internationella utvecklingen och ISO får du rätt kunskap i rätt tid, direkt från källan. Med vår kunskap om standarders möjligheter hjälper vi våra kunder att skapa verklig nytta och lönsamhet i sina verksamheter.

Vill du veta mer om SIS eller hur standarder kan effektivisera din verksamhet är du välkommen in på www.sis.se eller ta kontakt med oss på tel 08-555 523 00.



Standards make the world go round

SIS (Swedish Standards Institute) is an independent non-profit organisation with members from both the private and public sectors. We are part of the European and global network that draws up international standards. Standards consist of documented knowledge developed by prominent actors within the industry, business world and society. They promote cross-border trade, they help to make processes and products safer and they streamline your organisation.

Take part and have influence

As a member of SIS you will have the possibility to participate in standardization activities on national, European and global level. The membership in SIS will give you the opportunity to influence future standards and gain access to early stage information about developments within your field.

Get to know the finished work

We offer our customers everything in connection with standards and their application. You can purchase all the publications you need from us - everything from individual standards, technical reports and standard packages through to manuals and online services. Our web service e-nav gives you access to an easy-to-navigate library where all standards that are relevant to your company are available. Standards and manuals are sources of knowledge. We sell them.

Increase understanding and improve perception

With SIS you can undergo either shared or in-house training in the content and application of standards. Thanks to our proximity to international development and ISO you receive the right knowledge at the right time, direct from the source. With our knowledge about the potential of standards, we assist our customers in creating tangible benefit and profitability in their organisations.

If you want to know more about SIS, or how standards can streamline your organisation, please visit www.sis.se or contact us on phone +46 (0)8-555 523 00



Europastandarden EN 16280:2012 gäller som svensk standard. Detta dokument innehåller den officiella engelska versionen av EN 16280:2012.

The European Standard EN 16280:2012 has the status of a Swedish Standard. This document contains the official version of EN 16280:2012.

© Copyright/Upphovsrätten till denna produkt tillhör SIS, Swedish Standards Institute, Stockholm, Sverige. Användningen av denna produkt regleras av slutanvändarlicensen som återfinns i denna produkt, se standardens sista sidor.

© Copyright SIS, Swedish Standards Institute, Stockholm, Sweden. All rights reserved. The use of this product is governed by the end-user licence for this product. You will find the licence in the end of this document.

Upplysningar om sakinnehållet i standarden lämnas av SIS, Swedish Standards Institute, telefon 08-555 520 00. Standarder kan beställas hos SIS Förlag AB som även lämnar allmänna upplysningar om svensk och utländsk standard.

Information about the content of the standard is available from the Swedish Standards Institute (SIS), telephone +46 8 555 520 00. Standards may be ordered from SIS Förlag AB, who can also provide general information about Swedish and foreign standards.

Denna standard är framtagen av kommittén för Alkoholmätare, SIS/TK 486.

Har du synpunkter på innehållet i den här standarden, vill du delta i ett kommande revideringsarbete eller vara med och ta fram andra standarder inom området? Gå in på www.sis.se - där hittar du mer information.

EUROPEAN STANDARD

EN 16280

NORME EUROPÉENNE

EUROPÄISCHE NORM

October 2012

ICS 13.320; 71.040.40

English Version

Breath alcohol test devices for general public - Requirements and test methods

Ethylotests pour le grand public - Exigences et méthodes d'essais

Atemalkoholtestgeräte für den allgemeinen Gebrauch

This European Standard was approved by CEN on 18 August 2012.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents

Page

Foreword.....	3
Introduction	4
1 Scope	5
2 Normative references	5
3 Terms and definitions	5
4 Type-testing.....	6
5 Safety	6
5.1 General comments.....	6
5.2 Hygiene	7
5.3 Electrical safety.....	7
6 General specifications	7
6.1 General requirements.....	7
6.2 Maximum permissible error (MPE).....	7
6.3 Measurement range	7
6.4 Operating environmental conditions	7
6.5 Ease of use	8
6.6 Breath sampling method.....	8
6.7 Expression of results	8
6.8 Adjustment	9
6.9 Calibration period	9
6.10 Start-up time.....	9
6.11 Time for accepting a specimen	9
6.12 Frequency of measurement.....	9
6.13 Power supply duration	10
7 Metrological characteristics for testing	10
7.1 General conditions	10
7.2 Test gas characteristics	10
7.3 Reference conditions	11
7.4 Accuracy tests	11
7.5 Memory effect.....	12
7.6 Influence factors	12
7.7 Mechanical and climatic disturbances	14
8 Marking	16
9 Operating instructions	16
Annex A (informative) Example of type testing requirements.....	18
A.1 Additional information for compliance testing.....	18
A.2 Certification procedure	18
Annex B (informative) Figures for format 1 and format 2 - 6.7.4 Display (in normal mode).....	19
Annex C (informative) Formula for wet gas alcohol concentration	21
Bibliography	22

Foreword

This document (EN 16280:2012) has been prepared by Technical Committee CEN/TC 367 “Breath-alcohol testers”, the secretariat of which is held by AFNOR.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by April 2013, and conflicting national standards shall be withdrawn at the latest by April 2013.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

According to the CEN/CENELEC Internal Regulations, the national standards organisations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Introduction

The two main objectives of breath testers, the specifications of which are defined in this document, consist, on the one hand, in contributing to the prevention of accidents related to the consumption of alcohol, in particular road accidents, and on the other hand, in educating and making consumers aware of their responsibilities by enabling them to measure their breath alcohol level.

The purpose of this standard is to define requirements for a device which is capable of producing measurements that will deter a person who has consumed alcohol from driving or carrying out other risk-related activities.

Any appropriate technology capable of providing the functionality required in this document may be used.

The intention of the standard is to define specifications for a breath alcohol tester which will benefit to the general public at an affordable level.

The requirements in this standard apply for electronic devices only.

1 Scope

This European Standard applies to breath alcohol test devices which measure the concentration of alcohol contained in an exhaled breath sample, designed and intended to be used as a self tester for the general public and to provide a reliable indication of the breath alcohol concentration at the time of the test.

This European Standard specifies requirements for basic safety and performance, test methods and requirements for marking, labelling and operating instructions.

This European Standard gives guidelines for compliance testing procedures consisting of a number of technical performance tests.

It is not intended that the results of these devices should be used to rebut the results of evidential breath alcohol analysers covered by OIML R 126:1998, or breath alcohol test devices used in professional applications covered by EN 15964 or similar national regulations.

Therefore, the results of measurements need to be displayed so as to protect, as far as it is practicable, the user from underestimating his alcohol concentration based on measurement uncertainties, intrinsic in every measurement.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 60068-2-1, *Environmental testing — Part 2-1: Tests — Test A: Cold (IEC 60068-2-1)*

EN 60068-2-2, *Environmental testing — Part 2-2: Tests — Test B: Dry heat (IEC 60068-2-2)*

EN 60068-2-30, *Environmental testing — Part 2-30: Tests — Test Db: Damp heat, cyclic (12 h + 12 h cycle) (IEC 60068-2-30)*

EN 60068-2-32, *Basic environmental testing procedures — Part 2: Tests — Test Ed: Free fall (IEC 60068-2-32)*

EN 60335-2-29, *Household and similar electrical appliances — Safety — Part 2-29: Particular requirements for battery chargers (IEC 60335-2-29)*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1

alcohol

considered to be ethanol

3.2

breath alcohol test device

device which accepts a breath specimen, measures the concentration and indicates the level of alcohol in that breath specimen

3.3

operating state

state of the device in which it is able to take a breath specimen and determine the alcohol level in that breath specimen

SS-EN 16280:2012 (E)

3.4 normal mode
mode in which the device is ready to measure and display the level of alcohol in the breath specimen of the subject under test, either quantitatively or by preset level indication

3.5 test mode
mode in which the device displays the result of a test specified in this standard expressed in numerical format

3.6 unit of measurement
concentration of ethanol expressed in milligrams of ethanol per litre of exhaled volume (mg/L)

Note 1 to entry: Concentration in ethanol may be expressed in any other equivalent units, e.g. µg/L or µg/100 ml.

3.7 manufacturer
natural or legal person with responsibility for the design, manufacture, packaging and labelling of a device before it is placed on the market under his own name, regardless of whether these operations are carried out by that person himself or on his behalf by a third party

3.8 Maximum Permissible Error
MPE
extreme allowed value of measurement with respect to a test gas concentration defined in this standard

3.9 adjustment
process required to correct the measurement value of the breath alcohol test device when it is found to be outside of the defined MPE

3.10 calibration
process required to establish a relation between the measurement value of the breath alcohol test device and a reference gas

3.11 mouthpiece
disposable hygienically wrapped part that is fitted to the breath alcohol test device through which the subject under test provides the breath specimen, and that is used to prevent the breath sample being mixed with ambient air and diluting the alcohol concentration

4 Type-testing

An example of type-testing requirements is described in Annex A (informative).

5 Safety

5.1 General comments

The device shall be designed as far as possible to ensure the safety of the user of the device. Particular attention shall be made to the design and use of electrical connections as well as the materials chosen for mouthpiece construction and packaging.

5.2 Hygiene

The device shall be capable of use under hygienic conditions. It shall preclude the possibility of inhaling contaminated air from previous usages. It shall be possible to insert and remove these mouthpieces without touching the part which will be in contact with the lips of the user. The mouthpieces shall be supplied in individual, easily opened sealed packaging.

5.3 Electrical safety

The device shall be capable of operating within the requirements of relevant electrical safety regulations and standards. A battery charger or an external power supply provided as an accessory to the device shall be compliant with EN 60335-2-29.

6 General specifications

6.1 General requirements

It shall be clearly apparent when the device is ready to accept a breath specimen.

The device may have provision for manual acceptance of the test gas presented to it when conducting adjustment or calibration operations as well as metrological tests.

Devices shall be provided with an indication when the internal power supply is becoming exhausted.

If this low power indication is given, either the device stops operating or the device shall be capable of running further measurements according to the MPEs. The battery warning indicator shall not lead to confusion with any other displayed function.

Devices that also use external power supply shall be provided with an indicator that displays that power is on. This indicator shall not lead to confusion with any other displayed function.

The means by which the device is calibrated or adjusted shall only be accessible to authorised persons.

General device functions shall be verified to ensure that the device performs in accordance with the manufacturer's information.

6.2 Maximum permissible error (MPE)

The maximum permissible error is 0,04 mg/L for nominal alcohol concentration up to and including 0,20 mg/L.

The maximum permissible error is 20 % of nominal alcohol concentration above 0,20 mg/L.

6.3 Measurement range

Devices shall be capable of measuring alcohol concentrations according to the MPE in the range going from 0,00 to 0,50 mg/L.

6.4 Operating environmental conditions

— Temperature:

The devices shall be capable of use between + 10 °C and + 40 °C.

If the manufacturer specifies that the device may be operated outside this range, then it shall fulfil the requirements of this standard for these conditions.

If the device is operated outside the temperature range, then this shall be indicated.