

SVENSK STANDARD

SS-EN 16242:2012



Fastställt/Approved: 2012-12-14

Publicerad/Published: 2012-12-17 (Korrigerad version/Corrected version, January 2013)


Utgåva/Edition: 1

Språk/Language: engelska/English

ICS: 97.195

Bevarande av kulturarv – Metoder och instrument för att mäta relativ fuktighet i luft och fuktutbyte mellan luft och material

Conservation of cultural heritage – Procedures and instruments for measuring humidity in the air and moisture exchanges between air and cultural property



Standarder får världen att fungera

SIS (Swedish Standards Institute) är en fristående ideell förening med medlemmar från både privat och offentlig sektor. Vi är en del av det europeiska och globala nätverk som utarbetar internationella standarder. Standarder är dokumenterad kunskap utvecklad av framstående aktörer inom industri, näringsliv och samhälle och befrämjar handel över gränser, bidrar till att processer och produkter blir säkrare samt effektiviserar din verksamhet.

Delta och påverka

Som medlem i SIS har du möjlighet att påverka framtida standarder inom ditt område på nationell, europeisk och global nivå. Du får samtidigt tillgång till tidig information om utvecklingen inom din bransch.

Ta del av det färdiga arbetet

Vi erbjuder våra kunder allt som rör standarder och deras tillämpning. Hos oss kan du köpa alla publikationer du behöver – allt från enskilda standarder, tekniska rapporter och standardpaket till handböcker och onlinetjänster. Genom vår webbtjänst e-nav får du tillgång till ett lättnavigerat bibliotek där alla standarder som är aktuella för ditt företag finns tillgängliga. Standarder och handböcker är källor till kunskap. Vi säljer dem.

Utveckla din kompetens och lyckas bättre i ditt arbete

Hos SIS kan du gå öppna eller företagsinterna utbildningar kring innehåll och tillämpning av standarder. Genom vår närhet till den internationella utvecklingen och ISO får du rätt kunskap i rätt tid, direkt från källan. Med vår kunskap om standarders möjligheter hjälper vi våra kunder att skapa verklig nytta och lönsamhet i sina verksamheter.

Vill du veta mer om SIS eller hur standarder kan effektivisera din verksamhet är du välkommen in på www.sis.se eller ta kontakt med oss på tel 08-555 523 00.



Standards make the world go round

SIS (Swedish Standards Institute) is an independent non-profit organisation with members from both the private and public sectors. We are part of the European and global network that draws up international standards. Standards consist of documented knowledge developed by prominent actors within the industry, business world and society. They promote cross-border trade, they help to make processes and products safer and they streamline your organisation.

Take part and have influence

As a member of SIS you will have the possibility to participate in standardization activities on national, European and global level. The membership in SIS will give you the opportunity to influence future standards and gain access to early stage information about developments within your field.


Get to know the finished work

We offer our customers everything in connection with standards and their application. You can purchase all the publications you need from us - everything from individual standards, technical reports and standard packages through to manuals and online services. Our web service e-nav gives you access to an easy-to-navigate library where all standards that are relevant to your company are available. Standards and manuals are sources of knowledge. We sell them.

Increase understanding and improve perception

With SIS you can undergo either shared or in-house training in the content and application of standards. Thanks to our proximity to international development and ISO you receive the right knowledge at the right time, direct from the source. With our knowledge about the potential of standards, we assist our customers in creating tangible benefit and profitability in their organisations.

If you want to know more about SIS, or how standards can streamline your organisation, please visit www.sis.se or contact us on phone +46 (0)8-555 523 00



Europastandarden EN 16242:2012 gäller som svensk standard. Detta dokument innehåller den officiella engelska versionen av EN 16242:2012.

The European Standard EN 16242:2012 has the status of a Swedish Standard. This document contains the official version of EN 16242:2012.

**Denna korrigerade version innehåller följande ändring /
This corrected version contains the following corrections:**

Annex A (Clause A.3) the formulae A.12 and A.13 have been corrected.

Annex C (Clause C.1) the key to figure C.1 has been corrected.

© Copyright/Upphovsrätten till denna produkt tillhör SIS, Swedish Standards Institute, Stockholm, Sverige. Användningen av denna produkt regleras av slutanvändarlicensen som återfinns i denna produkt, se standardens sista sidor.

© Copyright SIS, Swedish Standards Institute, Stockholm, Sweden. All rights reserved. The use of this product is governed by the end-user licence for this product. You will find the licence in the end of this document.

Uppllysningar om sakinnehållet i standarden lämnas av SIS, Swedish Standards Institute, telefon 08-555 520 00. Standarder kan beställas hos SIS Förlag AB som även lämnar allmänna uppllysningar om svensk och utländsk standard.

Information about the content of the standard is available from the Swedish Standards Institute (SIS), telephone +46 8 555 520 00. Standards may be ordered from SIS Förlag AB, who can also provide general information about Swedish and foreign standards.

Denna standard är framtagen av kommittén för Bevarande av kulturarv, SIS/TK 479.

Har du synpunkter på innehållet i den här standarden, vill du delta i ett kommande revideringsarbete eller vara med och ta fram andra standarder inom området? Gå in på www.sis.se - där hittar du mer information.

EUROPEAN STANDARD

EN 16242

NORME EUROPÉENNE

EUROPÄISCHE NORM

November 2012

ICS 97.195

English Version

**Conservation of cultural heritage - Procedures and instruments
for measuring humidity in the air and moisture exchanges
between air and cultural property**

Conservation des biens culturels - Modes opératoires et
instruments de mesure de l'humidité de l'air et des
échanges d'humidité entre l'air et les biens culturels

Erhaltung des kulturellen Erbes - Verfahren und Geräte zur
Messung der Luftfeuchte und des Austausches von
Feuchtigkeit zwischen Luft und Kulturgut

This European Standard was approved by CEN on 8 September 2012.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents	Page
Foreword.....	3
Introduction	4
1 Scope	5
2 Normative references	5
3 Terms and definitions	5
4 Quantities characterising humidity in air	8
5 Considerations and recommendations related to measuring methods	9
5.1 Considerations.....	9
5.2 Recommendations.....	10
6 Main features of the hygrometers	11
6.1 Chilled-mirror dew-point hygrometer	11
6.2 Electronic psychrometer.....	12
6.3 Electronic hygrometer with a capacitive sensor	13
6.4 Electronic hygrometer with a resistive sensor.....	13
6.5 Hair hygrometer/hygrograph	14
7 Instrument calibration	14
Annex A (informative) Formulae for calculating relative humidity and related variables	16
A.1 Instruments: Psychrometer, barometer – Parameters: air temperature t ($^{\circ}\text{C}$), wet bulb air temperature t_w ($^{\circ}\text{C}$), p (hPa)	16
A.2 Instruments: RH hygrometer, thermometer, barometer - Parameters: t , RH , p	17
A.3 Instruments: Dew-point hygrometer, thermometer, barometer - Parameters: t , t_d , p	18
Annex B (informative) Examples for indoor climate measurements	19
B.1 Recognising the penetration and spread of external air across a room	19
B.2 Recognising if wall dampness is associated to condensation or evaporation	20
B.3 External dampness entering a room shown with a mixing ratio plot.....	20
Annex C (informative) Instrumental errors	22
C.1 Psychrometer: errors in the various hygrometric variables generated by an error of $0,1^{\circ}\text{C}$ in a temperature reading	22
C.2 Psychrometer: error in determining the relative humidity due to pressure change	23
C.3 Error due to a thermal inertia of a case, a probe or a shield.....	23
C.4 Typical non-linearity and hysteresis of the hair hygrometer	24
C.4.1 Hair non-linearity and hysteresis	24
C.4.2 Linear and non-linear scales	25
Bibliography	28

Foreword

This document (EN 16242:2012) has been prepared by Technical Committee CEN/TC 346 “Conservation of cultural heritage”, the secretariat of which is held by UNI.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by May 2013, and conflicting national standards shall be withdrawn at the latest by May 2013.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

According to the CEN/CENELEC Internal Regulations, the national standards organisations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Introduction

Humidity plays a key role in the conservation of cultural heritage because most materials and/or deterioration mechanisms are directly or indirectly affected by humidity levels or changes. This European Standard is a guide intended to assist in providing an acceptable environment for cultural heritage objects. Humidity in air, expressed in a number of ways, is an important aspect of that environment. Therefore, the control of levels and variability of humidity reduces the risk of deterioration and is an important preventive measure, minimising the need for future conservation interventions.

This European Standard is a guide to specifying adequate procedures for measuring humidity in air and the minimum characteristics of instruments for such measurements so that they are carried out to an appropriate level of accuracy. Although standards exist for measuring humidity in air in other fields like meteorology or ergonomics of thermal environments, this standard focuses on the specific requirements of cultural objects.

This document is one of the series of European Standards intended for use in the study of environments for cultural property.

1 Scope

This European Standard gives guidance and specifies procedures and instruments for the measurement of relative humidity (RH) in air, in outdoor or indoor environments. It indicates how RH can be directly measured or how it can be calculated from air temperature, wet-bulb temperature and dew-point temperature. This standard contains recommendations for accurate measurements of ambient conditions and moisture exchanges between air and cultural heritage objects. It is addressed to anyone in charge of environmental diagnosis, conservation or maintenance of buildings, collections or single objects.

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 15757:2010, *Conservation of Cultural Property - Specifications for temperature and relative humidity to limit climate-induced mechanical damage in organic hygroscopic materials*

EN 15758:2010, *Conservation of Cultural Property - Procedures and instruments for measuring temperatures of the air and of the surfaces of objects*

EN 60751, *Industrial platinum resistance thermometers and platinum temperature sensors (IEC 60751)*

EN ISO/IEC 17025, *General requirements for the competence of testing and calibration laboratories (ISO/IEC 17025)*

ISO/IEC Guide 98-3 *Uncertainty of measurement -- Part 3: Guide to the expression of uncertainty in measurement (GUM)*

3 Terms and definitions

For the purposes of this document, the terms and definitions given in EN 15758:2010 and the following apply.

3.1

absolute humidity (AH)

volume density of water vapour, i.e. the mass of vapour contained in the unit volume of moist air $AH = \frac{m_v}{V}$, expressed in g/m³

Note 1 to entry: This volume density is also noted ρ_v (v for volume)

3.2

atmospheric (or barometric) pressure (p)

pressure is the force per unit area exerted by the air column above the measuring point, expressed in hPa (hectopascal)

Note 1 to entry: 1 hPa = 1 mbar (millibar)

3.3

barometer

instrument for measuring atmospheric pressure

SS-EN 16242:2012 (E)

3.4
dew-point hygrometer
 instrument for measuring the temperature at which a cooled parcel of air becomes saturated with water vapour

3.5
dew-point temperature (DP)
 temperature to which air is cooled at constant pressure and constant water vapour content in order for saturation to occur

Note 1 to entry: This is expressed in degrees Celsius (°C). [EN 15758: 2010]

3.6
dry air
 atmospheric air without water vapour

3.7
dry-bulb temperature (T, t)
 actual air temperature. In a *psychrometer*, the temperature reached by the thermometer having the dry bulb

Note 1 to entry: Capital *T* is used when the measurement is expressed in Kelvin (K); lowercase *t* when expressed in degrees Celsius (°C).

3.8
equilibrium moisture content (EMC)
 moisture content at which a material neither loses nor gains moisture from the surrounding atmosphere at given relative humidity and temperature levels. Expressed in g/kg as the ratio of the mass of water m_w contained in the material and the dry mass m_{dm} of the same material, i.e.: $EMC = \frac{m_w}{m_{dm}}$

3.9
frost-point temperature
 temperature to which moist air must be cooled, at constant pressure and humidity mixing ratio, in order that it shall be saturated with respect to ice, expressed in degrees Celsius (°C)

3.10
hygrograph
 instrument for measuring relative humidity (see hygrometer) and recording over time

Note 1 to entry: Generally, a mechanically or electrically driven drum supporting a strip chart with RH graduation where an ink pen traces a time plot of the ambient humidity.

3.11
hygrometer
 instrument measuring relative humidity

Note 1 to entry: It generally comprises a sensor, which is set in equilibrium with the air, and a system that transforms the signal from the sensor into humidity readings.

3.12
mixing ratio or humidity mixing ratio (MR)
 ratio of the mass of water vapour m_v to the mass of dry air m_a , i.e. $MR = \frac{m_v}{m_a}$, expressed in g/kg

3.13
moist air
 mixture of dry air and water vapour

3.14**psychrometer**

instrument for measuring the dry- and wet-bulb temperatures to calculate relative humidity and other related variables

Note 1 to entry: It consists of two identical thermometers, one of which is sheathed in wet wicking, and a fan to ensure their ventilation at a constant velocity in order to reach equilibrium with air. Thermometer readings are expressed in degrees Celsius (°C). Some electronic instruments provide readings of relative humidity (%), dew point (°C) and other related variables.

3.15**relative humidity (RH)**

ratio of the actual vapour pressure of the air to the saturation vapour pressure

[SOURCE: EN 15757:2010]

3.16**measuring range**

interval of values that are intended to be measured, or that are potentially measurable, or that have been measured, specified by their upper and lower limits

3.17**repeatability**

ability of the measuring instrument to reproduce the same output when successively measuring the same value of the air or the surface under investigation, taken under the same conditions

Note 1 to entry: This is expressed as ± percent of the range.

[SOURCE: EN 15758:2010]

3.18**resolution**

smallest difference between indications of a displaying device that can be meaningfully distinguished

3.19**response time**

time interval between the instant when the parameter under investigation is subjected to a specified abrupt change and the instant when the response reaches and remains within specified limits around its final steady value

Note 1 to entry: The response time is typically expressed as the time needed to reach 63,2 % of the final value and in this case is called time constant, or 90 % or 95 % of it. The 90 % response time is 2,3 times longer than the time constant and the 95 % response time is three times longer. The response time is independent of the span of the output change.

[SOURCE: EN 15758:2010]

3.20**saturation vapour pressure ($e_{sat}(t)$)**

maximum pressure of the water vapour in equilibrium with plane surface of pure water, expressed in hPa (hectopascal)

3.21**sensor**

device that senses either an absolute value or a change in a physical quantity and converts them into a useful signal for an information-gathering system