

SVENSK STANDARD

SS-EN ISO 23908:2013



Fastställt/Approved: 2013-08-09
Publicerad/Published: 2013-08-15
Utgåva/Edition: 2
Språk/Language: engelska/English
ICS: 11.040.25

**Skydd mot stick och skärskador – Krav och testmetoder –
Skyddsfunktioner för stickande och skärande
engångsprodukter av typ injektionsnålar, introducers för
katetrar och nålar för blodprovstagning (ISO 23908:2011)**

**Sharps injury protection – Requirements and test methods –
Sharps protection features for single-use hypodermic needles,
introducers for catheters and needles used for blood sampling
(ISO 23908:2011)**

This preview is downloaded from www.sis.se. Buy the entire standard via <https://www.sis.se/std-98727>

Standarder får världen att fungera

SIS (Swedish Standards Institute) är en fristående ideell förening med medlemmar från både privat och offentlig sektor. Vi är en del av det europeiska och globala nätverk som utarbetar internationella standarder. Standarder är dokumenterad kunskap utvecklad av framstående aktörer inom industri, näringsliv och samhälle och befrämjar handel över gränser, bidrar till att processer och produkter blir säkrare samt effektiviserar din verksamhet.

Delta och påverka

Som medlem i SIS har du möjlighet att påverka framtida standarder inom ditt område på nationell, europeisk och global nivå. Du får samtidigt tillgång till tidig information om utvecklingen inom din bransch.



Ta del av det färdiga arbetet

Vi erbjuder våra kunder allt som rör standarder och deras tillämpning. Hos oss kan du köpa alla publikationer du behöver – allt från enskilda standarder, tekniska rapporter och standardpaket till handböcker och onlinetjänster. Genom vår webbtjänst e-nav får du tillgång till ett lättnavigerat bibliotek där alla standarder som är aktuella för ditt företag finns tillgängliga. Standarder och handböcker är källor till kunskap. Vi säljer dem.

Utveckla din kompetens och lyckas bättre i ditt arbete

Hos SIS kan du gå öppna eller företagsinterna utbildningar kring innehåll och tillämpning av standarder. Genom vår närhet till den internationella utvecklingen och ISO får du rätt kunskap i rätt tid, direkt från källan. Med vår kunskap om standarders möjligheter hjälper vi våra kunder att skapa verklig nytta och lönsamhet i sina verksamheter.

Vill du veta mer om SIS eller hur standarder kan effektivisera din verksamhet är du välkommen in på www.sis.se eller ta kontakt med oss på tel 08-555 523 00.



Standards make the world go round

SIS (Swedish Standards Institute) is an independent non-profit organisation with members from both the private and public sectors. We are part of the European and global network that draws up international standards. Standards consist of documented knowledge developed by prominent actors within the industry, business world and society. They promote cross-border trade, they help to make processes and products safer and they streamline your organisation.

Take part and have influence

As a member of SIS you will have the possibility to participate in standardization activities on national, European and global level. The membership in SIS will give you the opportunity to influence future standards and gain access to early stage information about developments within your field.

Get to know the finished work

We offer our customers everything in connection with standards and their application. You can purchase all the publications you need from us - everything from individual standards, technical reports and standard packages through to manuals and online services. Our web service e-nav gives you access to an easy-to-navigate library where all standards that are relevant to your company are available. Standards and manuals are sources of knowledge. We sell them.

Increase understanding and improve perception

With SIS you can undergo either shared or in-house training in the content and application of standards. Thanks to our proximity to international development and ISO you receive the right knowledge at the right time, direct from the source. With our knowledge about the potential of standards, we assist our customers in creating tangible benefit and profitability in their organisations.

If you want to know more about SIS, or how standards can streamline your organisation, please visit www.sis.se or contact us on phone +46 (0)8-555 523 00



Europastandarden EN ISO 23908:2013 gäller som svensk standard. Detta dokument innehåller den officiella engelska versionen av EN ISO 23908:2013.

Denna standard ersätter SS-EN ISO 23908:2011, utgåva 1.

The European Standard EN ISO 23908:2013 has the status of a Swedish Standard. This document contains the official version of EN ISO 23908:2013.

This standard supersedes the Swedish Standard SS-EN ISO 23908:2011, edition 1.

© Copyright/Upphovsrätten till denna produkt tillhör SIS, Swedish Standards Institute, Stockholm, Sverige. Användningen av denna produkt regleras av slutanvändarlicensen som återfinns i denna produkt, se standardens sista sidor.

© Copyright SIS, Swedish Standards Institute, Stockholm, Sweden. All rights reserved. The use of this product is governed by the end-user licence for this product. You will find the licence in the end of this document.

Upplysningar om sakinnehållet i standarden lämnas av SIS, Swedish Standards Institute, telefon 08-555 520 00. Standarder kan beställas hos SIS Förlag AB som även lämnar allmänna upplysningar om svensk och utländsk standard.

Information about the content of the standard is available from the Swedish Standards Institute (SIS), telephone +46 8 555 520 00. Standards may be ordered from SIS Förlag AB, who can also provide general information about Swedish and foreign standards.

Denna standard är framtagen av kommittén för Förbrukningsmaterial inom sjukvården, SIS/TK 330.

Har du synpunkter på innehållet i den här standarden, vill du delta i ett kommande revideringsarbete eller vara med och ta fram andra standarder inom området? Gå in på www.sis.se - där hittar du mer information.

EUROPEAN STANDARD

EN ISO 23908

NORME EUROPÉENNE

EUROPÄISCHE NORM

February 2013

ICS 11.040.25; 11.040.99

Supersedes EN ISO 23908:2011

English Version

**Sharps injury protection - Requirements and test methods -
Sharps protection features for single-use hypodermic needles,
introducers for catheters and needles used for blood sampling
(ISO 23908:2011)**

Protection contre les blessures par perforants - Exigences
et méthodes d'essai - Dispositifs de protection des aiguilles
hypodermiques, des introducteurs pour cathéters et des
aiguilles utilisées pour les prélèvements sanguins, non
réutilisables (ISO 23908:2011)

Schutz vor Stich- und Schnittverletzung - Anforderungen
und Prüfverfahren - Schutzeinrichtungen für einmalig zu
verwendende Nadeln zur subkutanen Injektion,
Kathetereinführungen und Nadeln zur Blutentnahme (ISO
23908:2011)

This European Standard was approved by CEN on 8 January 2013.

CEN members are bound to comply with the CEN/CENELEC Internal Regulations which stipulate the conditions for giving this European Standard the status of a national standard without any alteration. Up-to-date lists and bibliographical references concerning such national standards may be obtained on application to the CEN-CENELEC Management Centre or to any CEN member.

This European Standard exists in three official versions (English, French, German). A version in any other language made by translation under the responsibility of a CEN member into its own language and notified to the CEN-CENELEC Management Centre has the same status as the official versions.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents

Page

Foreword	iv
Introduction.....	v
1 Scope	1
2 Normative references	1
3 Terms and definitions	1
4 Requirements.....	2
4.1 General	2
4.2 Activation of the sharps injury protection feature	3
4.3 Security of safe mode protection	3
5 Test methods	3
5.1 General	3
5.2 Testing activation of a sharps injury protection feature	3
5.2.1 Principle.....	3
5.2.2 Apparatus.....	4
5.2.3 Procedure.....	4
5.3 Challenging the device in safe mode	5
5.3.1 General	5
5.3.2 Principle.....	5
5.3.3 Apparatus.....	5
5.3.4 Procedure.....	5
5.4 Testing access to the sharp in safe mode	6
5.5 Testing simulated clinical use	6
5.6 Test report.....	6
6 Information supplied by the manufacturer	6
6.1 General	6
6.2 Marking/labelling	7
6.3 Instructions for use	7
Annex A (informative) Guidance on simulated user studies	8
Annex B (informative) Method for testing access to the sharp in safe mode	10
Annex ZA (informative) Relationship between this European Standard and the Essential Requirements of EU Directive 93/42/EEC on Medical Devices.....	11
Bibliography.....	12

Foreword

The text of ISO 23908:2011 has been prepared by Technical Committee ISO/TC 84 “Devices for administration of medicinal products and intravascular catheters” of the International Organization for Standardization (ISO) and has been taken over as EN ISO 23908:2013 by Technical Committee CEN/TC 205 “Non-active medical devices” the secretariat of which is held by DIN.

This European Standard shall be given the status of a national standard, either by publication of an identical text or by endorsement, at the latest by August 2013, and conflicting national standards shall be withdrawn at the latest by August 2013.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document supersedes EN ISO 23908:2011.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive.

For relationship with EU Directive, see informative Annex ZA, which is an integral part of this document.

According to the CEN-CENELEC Internal Regulations, the national standards organizations of the following countries are bound to implement this European Standard: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Endorsement notice

The text of ISO 23908:2011 has been approved by CEN as EN ISO 23908:2013 without any modification.

SS-EN ISO 23908:2013 (E)

Introduction

This International Standard addresses sharps injury protection systems designed to protect users of medical devices. These sharps injury protection features are intended to prevent, or reduce the potential for, disease transmission which could result from accidental, post-use sharps injuries.

This International Standard is aimed at addressing devices primarily intended for human use, of a wide range of product types, including, among others, hollow-bore needles for injection or infusion of therapeutics into the body, or sampling of fluids from the body, and hollow bore or solid-core needles used for blood sampling (e.g. lancing devices). It addresses sharps injury protection systems which are either active or passive in their activation after the medical device's intended use. It does not cover solid-core needles used for surgery (e.g. suture needles).

Given the broad variation in product design and sharps protection technology, the variety of different types of devices, and in order to avoid unnecessarily restricting innovation, this International Standard has been developed as “horizontal” in nature, which means it provides for general design, testing and labelling requirements, rather than specific physical and prescriptive design requirements. It therefore differs from more “vertical” standards, which list specific maximum forces, detailed test fixture designs, test systems to be used or detailed test measures, as such prescriptive details cannot cover the variety of designs and devices, and may impede continuing innovation in new products, features and/or protection mechanisms that lead to future improvements in healthcare.

This International Standard presumes that the product developer would use a risk-based approach (consistent with ISO 14971) to determine the device design that best meets the needs of a target user population and expected use settings. Through this risk-based approach, the sharps injury protection system would have performance requirements appropriate to the foreseeable risks associated with the intended use of the device, expected user interfaces, and the settings in which these safety features are expected to be used.

This International Standard provides guidelines to enable the manufacturer to verify that the design of the sharps injury protection systems complies with the design intent spelled out in the design specification. As part of this verification, the manufacturer is expected to demonstrate that the performance of the sharps injury protection system is appropriate to the intended users and settings through the use of appropriate simulated or clinical use studies. These simulated or clinical use studies allow the manufacturer to demonstrate that, when used in accordance with the instructions for use, in settings representative of real-life intended use and by intended or foreseeable users, the device functions as intended.

Existing products and those currently under development may not fulfil some of the requirements given by this International Standard. However, manufacturers would be well advised to follow its provisions when improving existing products or developing new products to obtain an even higher level of quality.

Sharps injury protection — Requirements and test methods — Sharps protection features for single-use hypodermic needles, introducers for catheters and needles used for blood sampling

1 Scope

This International Standard gives requirements and test methods for evaluating the performance parameters of sharps injury protection features, whether active or passive in design, for medical devices containing (sharp) hypodermic needles for single use, introducers for catheters and lancets, and other needles used in blood sampling. The sharps injury protection devices it covers may be provided integral to the device or combined with the device prior to use to achieve the sharps injury protection.

It does not give requirements for the storage and handling of the sharps protection before its intended use, or for the medical device itself.

2 Normative references

The following referenced documents are indispensable for the application of this document. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

ISO 2859 (all parts), *Sampling procedures for inspection by attributes*

ISO 3951 (all parts), *Sampling procedures for inspection by variables*

ISO 14971, *Medical devices — Application of risk management to medical devices*

ISO 16269-6, *Statistical interpretation of data — Part 6: Determination of statistical tolerance intervals*

3 Terms and definitions

For the purposes of this document, the following terms and definitions apply.

3.1

activation

deployment of the sharps protection mechanism

3.2

active safety feature

sharps protection feature that requires an additional step by the user to activate, separate from any action needed to perform the primary intended function of the device

3.3

accidental sharps injury

unintentional penetration into human tissue by the sharp after the intended use

SS-EN ISO 23908:2013 (E)

3.4
passive safety feature
sharps protection feature that does not require an additional step by the user to activate, separate from any action needed to perform the primary intended function of the device

3.5
safe mode
state of the device after activation of the safety feature

3.6
sharp
part of the device that can penetrate human tissue

3.7
sharps injury protection feature
feature that prevents accidental sharps injury

4 Requirements

4.1 General

4.1.1 Where the requirements do not specify forces for activation of the safety feature, the appropriate force shall be determined by using a risk-based approach in accordance with ISO 14971, supported by simulated user studies that mimic actual clinical use by using patient substitutes (e.g. instructional models) rather than actual patients. The study design should be based on statistical considerations and should have clear acceptance criteria. Guidance on conducting simulated user studies is outlined in Annex A.

4.1.2 Once in safe mode, the safety feature(s) of the device shall provide protection against accidental sharps injury until safe disposal of the sharp under expected conditions of use.

4.1.3 It shall be apparent to the user as to when the device is in safe mode.

Activation/safe mode shall be communicated to the user in a clear and unmistakable manner by either visual, tactile and/or audible means. If the manufacturer determines that the user environment requires a permanent indication of safe mode, then a visual indication shall be included.

4.1.4 Activation of the sharps protection feature shall permit the user's hand(s) to remain behind the exposed contaminated sharp.

Safety features may be operated either actively or passively. If active operation is required, one-handed operation is recommended.

4.1.5 The safety result shall

- not negatively affect the intended performance characteristics or proper disposal of the device,
- not impede or adversely affect the intended clinical performance of the device,
- resist inadvertent activation under expected conditions of use.

4.1.6 The performance of the safety feature as described in 4.1.2 to 4.1.5 shall be demonstrated through appropriate simulated or clinical use studies for the specified conditions indicated under the conditions of use.

NOTE 1 Appropriate simulated or clinical use studies may be helpful in establishing specifications to meet the requirements of Clause 5.

NOTE 2 Annex A contains guidance for simulated or clinical use studies.

NOTE 3 IEC 62366 covers the application of usability engineering to medical devices.

4.2 Activation of the sharps injury protection feature

Active sharps injury protection features shall be able to be activated immediately after intended use.

Passive sharps injury protection features shall enter safe mode immediately after intended use.

The sharps injury protection feature shall be able to be activated by a force appropriate for the intended users of the device (e.g. patients, health care professionals or family members). An appropriate force shall be selected that eases actuation and avoids unintended actuation.

The appropriate activation forces shall be determined using a risk-based approach in accordance with ISO 14971. The manufacturer shall confirm that these force values are the values at which the sharps injury protection feature can be activated. These force values shall be obtained using the methodology outlined in Clause 5.

4.3 Security of safe mode protection

Once in safe mode, the safety feature shall

- a) resist forces so as to prevent unintended exposure to the sharps, when tested in accordance with 5.3, and
- b) minimize the risk of accidental access to the sharp, when tested in accordance with 5.4.

Using a risk-based approach in accordance with ISO 14971, the manufacturer shall determine appropriate minimum overriding forces. These force values shall be obtained using the methodology outlined in Clause 5.

5 Test methods

5.1 General

Unless otherwise specified in the relevant device standard(s), all tests and test evaluations shall be performed at the following standard atmosphere conditions:

- temperature: $(23 \pm 5) ^\circ\text{C}$;
- relative humidity: $(50 \pm 25) \%$.

The device with integrated sharps injury protection or stand-alone sharps injury protection device that is tested shall have been subjected to storage for at least 4 h under these conditions immediately prior to testing/evaluation.

The repeatability and reproducibility of the test apparatus shall be no greater than 20 % of the allowed tolerance band for any given set of measurements.

When a sharps protection means is integral to a device covered by any other standard, or when combined with such a device prior to use, it shall be subjected to the same preconditioning requirements set out for the device by that other standard.

5.2 Testing activation of a sharps injury protection feature

5.2.1 Principle

Test pieces shall be chosen to test the intended means of activation by either tensile, compressive or torsional force applied directly (but smoothly) to the safety mechanism. The resulting activation forces/torques shall be recorded as specified in 5.6.