

Teknisk specifikation

SIS-CEN/TS 16175-1:2013

Publicerad/Published: 2013-09-12

Utgåva/Edition: 1

Språk/Language: engelska/English

ICS: 13.030.01; 13.080.10

Slam, behandlat bioavfall och mark – Bestämning av kvicksilver i kungsvatten och salpetersyrauppslutning – Del 1: Atomabsorptionsspektrometri med kall ånga (CV-AAS)

Sludge, treated biowaste and soil – Determination of mercury – Part 1: Cold-vapour atomic absorption spectrometry (CV-AAS)

This preview is downloaded from www.sis.se. Buy the entire standard via <https://www.sis.se/std-98812>

Standarder får världen att fungera

SIS (Swedish Standards Institute) är en fristående ideell förening med medlemmar från både privat och offentlig sektor. Vi är en del av det europeiska och globala nätverk som utarbetar internationella standarder. Standarder är dokumenterad kunskap utvecklad av framstående aktörer inom industri, näringsliv och samhälle och befrämjar handel över gränser, bidrar till att processer och produkter blir säkrare samt effektiviserar din verksamhet.

Delta och påverka

Som medlem i SIS har du möjlighet att påverka framtida standarder inom ditt område på nationell, europeisk och global nivå. Du får samtidigt tillgång till tidig information om utvecklingen inom din bransch.


Ta del av det färdiga arbetet

Vi erbjuder våra kunder allt som rör standarder och deras tillämpning. Hos oss kan du köpa alla publikationer du behöver – allt från enskilda standarder, tekniska rapporter och standardpaket till handböcker och onlinetjänster. Genom vår webbtjänst e-nav får du tillgång till ett lättnavigerat bibliotek där alla standarder som är aktuella för ditt företag finns tillgängliga. Standarder och handböcker är källor till kunskap. Vi säljer dem.

Utveckla din kompetens och lyckas bättre i ditt arbete

Hos SIS kan du gå öppna eller företagsinterna utbildningar kring innehåll och tillämpning av standarder. Genom vår närhet till den internationella utvecklingen och ISO får du rätt kunskap i rätt tid, direkt från källan. Med vår kunskap om standarders möjligheter hjälper vi våra kunder att skapa verklig nytta och lönsamhet i sina verksamheter.

Vill du veta mer om SIS eller hur standarder kan effektivisera din verksamhet är du välkommen in på www.sis.se eller ta kontakt med oss på tel 08-555 523 00.



Standards make the world go round

SIS (Swedish Standards Institute) is an independent non-profit organisation with members from both the private and public sectors. We are part of the European and global network that draws up international standards. Standards consist of documented knowledge developed by prominent actors within the industry, business world and society. They promote cross-border trade, they help to make processes and products safer and they streamline your organisation.

Take part and have influence

As a member of SIS you will have the possibility to participate in standardization activities on national, European and global level. The membership in SIS will give you the opportunity to influence future standards and gain access to early stage information about developments within your field.

Get to know the finished work

We offer our customers everything in connection with standards and their application. You can purchase all the publications you need from us - everything from individual standards, technical reports and standard packages through to manuals and online services. Our web service e-nav gives you access to an easy-to-navigate library where all standards that are relevant to your company are available. Standards and manuals are sources of knowledge. We sell them.

Increase understanding and improve perception

With SIS you can undergo either shared or in-house training in the content and application of standards. Thanks to our proximity to international development and ISO you receive the right knowledge at the right time, direct from the source. With our knowledge about the potential of standards, we assist our customers in creating tangible benefit and profitability in their organisations.

If you want to know more about SIS, or how standards can streamline your organisation, please visit www.sis.se or contact us on phone +46 (0)8-555 523 00



Denna tekniska specifikation är inte en svensk standard. Detta dokument innehåller den engelska språkversionen av CEN/TS 16175-1:2013.

This Technical Specification is not a Swedish Standard. This document contains the English version of CEN/TS 16175-1:2013.

© Copyright/Upphovsrätten till denna produkt tillhör SIS, Swedish Standards Institute, Stockholm, Sverige. Användningen av denna produkt regleras av slutanvändarlicensen som återfinns i denna produkt, se standardens sista sidor.

© Copyright SIS, Swedish Standards Institute, Stockholm, Sweden. All rights reserved. The use of this product is governed by the end-user licence for this product. You will find the licence in the end of this document.

Uppllysningar om sakinnehållet i detta dokument lämnas av SIS, Swedish Standards Institute, telefon 08-555 520 00. Standarder kan beställas hos SIS Förlag AB som även lämnar allmänna uppllysningar om nationell och internationell standard.

Information about the content of this document is available from the SIS, Swedish Standards Institute, telephone +46 8 555 520 00. Standards may be ordered from SIS Förlag AB, who can also provide general information about national and international standards.

Dokumentet är framtaget av kommittén för Karaktärisering av avfall, mark och slam, SIS/TK 535.

Har du synpunkter på innehållet i det här dokumentet, vill du delta i ett kommande revideringsarbete eller vara med och ta fram standarder inom området? Gå in på www.sis.se - där hittar du mer information.

TECHNICAL SPECIFICATION
SPÉCIFICATION TECHNIQUE
TECHNISCHE SPEZIFIKATION

CEN/TS 16175-1

January 2013

ICS 13.030.01; 13.080.10

English Version

**Sludge, treated biowaste and soil - Determination of mercury -
Part 1: Cold-vapour atomic absorption spectrometry (CV-AAS)**

Boues, biodéchets traités et sols - Détermination du
mercure - Partie 1: Spectrométrie d'absorption atomique de
vapeur froide (SAA-VP)

Schlamm, behandelter Bioabfall und Boden - Bestimmung
von Quecksilber - Teil 1: Kaltdampf-
Atomabsorptionsspektrometrie (CV-AAS)

This Technical Specification (CEN/TS) was approved by CEN on 16 June 2012 for provisional application.

The period of validity of this CEN/TS is limited initially to three years. After two years the members of CEN will be requested to submit their comments, particularly on the question whether the CEN/TS can be converted into a European Standard.

CEN members are required to announce the existence of this CEN/TS in the same way as for an EN and to make the CEN/TS available promptly at national level in an appropriate form. It is permissible to keep conflicting national standards in force (in parallel to the CEN/TS) until the final decision about the possible conversion of the CEN/TS into an EN is reached.

CEN members are the national standards bodies of Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and United Kingdom.



EUROPEAN COMMITTEE FOR STANDARDIZATION
COMITÉ EUROPÉEN DE NORMALISATION
EUROPÄISCHES KOMITEE FÜR NORMUNG

Management Centre: Avenue Marnix 17, B-1000 Brussels

Contents		Page
Foreword.....		3
Introduction		4
1 Scope		5
2 Normative references		5
3 Principle.....		5
4 Interferences		5
5 Reagents.....		6
6 Apparatus		7
7 Procedure		8
7.1 Test solution.....		8
7.2 Test blank solution		8
7.3 Preparation of the calibration solutions.....		8
7.4 Calibration		8
7.5 Measurement of test sample		8
8 Calculation and expression of results.....		9
8.1 Calculation.....		9
8.2 Expression of results		9
9 Precision data		9
10 Test report		10
Annex A (informative) Repeatability and reproducibility data.....		11
A.1 Materials used in the interlaboratory comparison study		11
A.2 Interlaboratory comparison results		12
Bibliography		13

Foreword

This document (CEN/TS 16175-1:2013) has been prepared by Technical Committee CEN/TC 400 "Project Committee - Horizontal standards in the fields of sludge, biowaste and soil", the secretariat of which is held by DIN.

Attention is drawn to the possibility that some of the elements of this document may be the subject of patent rights. CEN [and/or CENELEC] shall not be held responsible for identifying any or all such patent rights.

This document has been prepared under a mandate given to CEN by the European Commission and the European Free Trade Association, and supports essential requirements of EU Directive(s).

The preparation of this document by CEN is based on a mandate by the European Commission (Mandate M/330), which assigned the development of standards on sampling and analytical methods for hygienic and biological parameters as well as inorganic and organic determinants, aiming to make these standards applicable to sludge, treated biowaste and soil as far as this is technically feasible.

CEN/TS 16175, *Sludge, treated biowaste and soil — Determination of mercury*, is comprised of the following parts:

- *Part 1: Cold-vapour atomic absorption spectrometry (CV-AAS)*;
- *Part 2: Cold-vapour atomic fluorescence spectrometry (CV-AFS)*.

According to the CEN/CENELEC Internal Regulations, the national standards organisations of the following countries are bound to announce this Technical Specification: Austria, Belgium, Bulgaria, Croatia, Cyprus, Czech Republic, Denmark, Estonia, Finland, Former Yugoslav Republic of Macedonia, France, Germany, Greece, Hungary, Iceland, Ireland, Italy, Latvia, Lithuania, Luxembourg, Malta, Netherlands, Norway, Poland, Portugal, Romania, Slovakia, Slovenia, Spain, Sweden, Switzerland, Turkey and the United Kingdom.

Introduction

This Technical Specification is validated for several types of matrices as indicated in Table 1 (see also Annex A for the results of validation).

Table 1 — Matrices for which this Technical Specification is applicable and validated

Matrix	Materials used for validation
Sludge	Municipal sludge
Biowaste	Compost
Soil	Agricultural soil

WARNING — Persons using this Technical Specification should be familiar with usual laboratory practice. This Technical Specification does not purport to address all of the safety problems, if any, associated with its use. It is the responsibility of the user to establish appropriate safety and health practices and to ensure compliance with any national regulatory conditions.

IMPORTANT — It is absolutely essential that tests conducted according to this Technical Specification be carried out by suitably trained staff.

1 Scope

This Technical Specification specifies a method for the determination of mercury in *aqua regia* or nitric acid digests of sludge, treated biowaste and soil, obtained according to EN 16173 or EN 16174 using cold-vapour atomic absorption spectrometry (CV-AAS). The lower working range limit is 0,03 mg/kg (dry matter basis).

2 Normative references

The following documents, in whole or in part, are normatively referenced in this document and are indispensable for its application. For dated references, only the edition cited applies. For undated references, the latest edition of the referenced document (including any amendments) applies.

EN 15934, *Sludge, treated biowaste, soil and waste — Calculation of dry matter fraction after determination of dry residue or water content*

EN 16173, *Sludge, treated biowaste and soil — Digestion of nitric acid soluble fractions of elements*

EN 16174, *Sludge, treated biowaste and soil — Digestion of aqua regia soluble fractions of elements*

EN ISO 3696, *Water for analytical laboratory use — Specification and test methods (ISO 3696)*

3 Principle

Mono- and divalent mercury is reduced to the elemental form by tin(II) chloride or sodium borohydride in acid medium. Elemental mercury is stripped off from the solution in a closed system. The mercury vapour, in the form of an atomic gas, passes through a cell positioned in the light path of an atomic absorption spectrometer. Its absorbance at a wavelength of 253,7 nm is measured. The absorbance signal is a function of mercury concentration, and the concentrations are calculated using a calibration curve.

NOTE The matrix of the solution analysed is dominated by the acids used in the digestion step. Tin(II) chloride as a reduction substance is recommended in this Technical Specification, because sodium borohydride reduces many elements commonly found in soil, sludge and waste extract solutions, to the elemental state, which may cause matrix problems under particular circumstances. However, it is still possible to use sodium borohydride as reduction agent. The concentration range 1 µg/l to 10 µg/l in the digested solution, corresponding to 0,03 µg/g to 0,3 µg/g of mercury when a 3,0 g sample has been digested, can be determined directly. Higher concentrations can be determined if the digested solution is diluted. Sensitivity can be increased by the amalgamation technique.

4 Interferences

Because the samples are digested with oxidizing, strong acids before determination of mercury starts, there are no interferences with organic substances in the gas phase that can absorb in the UV range. Neither are there any interferences with mercury compounds which cannot be fully reduced with tin(II) chloride alone. The interferences due to the presence of other elements in the matrix depend on their concentrations. Element concentrations exceeding those listed in Table 2 may cause negative bias.

Fewer interferences arise from heavy metals when tin(II) chloride is used rather than sodium borohydride. When flow systems are used, interference effects due to heavy metals may be less than indicated in Table 2.

Table 2 — Tolerable concentrations of some matrix elements

Element	Acceptable concentration in the test solution mg/l
Cu(II)	500
Ni(II)	500
Ag(I)	1

5 Reagents

For the determination of mercury at trace and ultra-trace level, the reagents shall be of adequate purity. The concentration of mercury or interfering substances in the reagents and the water should be negligible compared to the lowest concentration to be determined.

5.1 Water quality 2 according to EN ISO 3696 for all sample preparations and dilutions.

5.2 Carrier gas, argon or nitrogen, with a purity of $\geq 99,99$ %.

5.3 Hydrochloric acid, HCl, $\rho(\text{HCl}) = 1,18$ g/ml, $c(\text{HCl}) = 12$ mol/l, $w(\text{HCl}) = 370$ g/kg.

The same batch of hydrochloric acid shall be used throughout the procedure.

5.4 Nitric acid, HNO_3 , $\rho(\text{HNO}_3) = 1,4$ g/ml, $c(\text{HNO}_3) = 15$ mol/l, $w(\text{HNO}_3) = 650$ g/kg.

The same batch of nitric acid shall be used throughout the procedure.

5.5 Nitric acid, diluted solution.

Add 10 ml nitric acid (5.4) to 500 ml of water in a 1 000 ml volumetric flask, mix and fill to the mark with water.

5.6 Nitric acid, $c(\text{HNO}_3) = 2$ mol/l, rinsing solution for glassware.

Add 150 ml of nitric acid (5.4) to about 500 ml of water, and dilute with water to 1 000 ml.

5.7 Aqua regia, diluted solution.

Add 21 ml hydrochloric acid (5.3) and 7 ml nitric acid (5.4) to 500 ml of water in a 1 000 ml volumetric flask, mix and fill to the mark with water.

5.8 Tin(II) chloride solution, $\rho(\text{SnCl}_2 \cdot 2 \text{H}_2\text{O}) = 100$ g/l.

Dissolve 10 g of $\text{SnCl}_2 \cdot 2 \text{H}_2\text{O}$ in 30 ml of hydrochloric acid (5.3), transfer to a 100 ml volumetric flask and fill to the mark with water. The blank concentration of mercury can be reduced by bubbling a stream of nitrogen through the solution for 30 min, if necessary. Prepare this solution on the day of use.

A solution of lower concentration, e. g. 0,5 g in 100 ml, may be used with flow systems. Prepare this latter solution freshly on the day of use from the more concentrated solution by diluting with water. Follow the recommendations of the manufacturer of the spectrometer.

5.9 Sodium borohydride solution, NaBH_4 , $\rho(\text{NaBH}_4) = 30 \text{ g/l}$.

1 g sodium hydroxide, NaOH , is weighed into a 100 ml volumetric flask and dissolved in water. 3 g sodium borohydride, NaBH_4 , are weighed into a 100 ml volumetric flask, dissolved and diluted to the mark with the sodium hydroxide solution.

A solution of lower concentration, e. g. 3 g/l may be used with flow systems. Prepare this latter solution freshly on the day of use from the more concentrated solution by diluting with water. Follow the recommendations of the manufacturer of the spectrometer.

5.10 Mercury standard stock solution, 1 000 mg/l.

Use a commercially available quantitative stock solution with a mercury concentration of $(1\ 000 \pm 2) \text{ mg/l}$. This solution is considered to be stable for at least one year, but in reference to guaranteed stability, see the recommendations of the manufacturer.

5.11 Mercury, standard solution I, 100 mg/l.

Pipette 10 ml of the stock mercury solution (5.10) into a 100 ml volumetric flask, add 10 ml nitric acid (5.4), mix and fill to the mark with water.

This solution is stable for one month.

5.12 Mercury, standard solution II, 1 mg/l.

Pipette 1 ml of the mercury standard solution I (5.11) into a 100 ml volumetric flask, add 10 ml nitric acid (5.4), mix and fill to the mark with water.

This solution is stable for 7 days.

5.13 Mercury, standard solution III, 100 $\mu\text{g/l}$.

Pipette 10 ml of the mercury standard solution II (5.12) into a 100 ml volumetric flask, add 10 ml nitric acid (5.4), mix and fill to the mark with water.

This solution shall be freshly prepared on the day of use.

6 Apparatus

6.1 Usual laboratory apparatus

All glassware shall be carefully cleaned for low trace element determinations, e. g. by immersion in nitric acid rinsing solution (5.6) for a minimum of 6 h, followed by rinsing with water before use. The nitric acid shall be replaced each week.

6.2 Atomic absorption spectrometer (AAS)

Instrument equipped with an appropriate monitoring system, with a cold-vapour generation system and a quartz or glass cuvette. As a radiation source use preferably an electrodeless discharge lamp (which gives a greater and more stable light intensity) operated at a current recommended by the lamp and instrument manufacturer, or a hollow cathode lamp. An AAS system with background correction device is recommended.

6.3 Automated sample introduction system

Automated mercury flow systems (flow injection systems or continuous flow systems) are very common for atomic absorption spectrometry. They allow a concentration range lower by about one order of magnitude. Manually operated systems or semi-automated batch systems are adequate as well.